

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081019\
 Data File : VN057164.D
 Acq On : 10 Aug 2019 9:50
 Operator : JC/SP
 Sample : VSTDIC005
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005

Quant Time: Aug 11 03:35:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N081019W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sun Aug 11 03:32:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.58	128	167661	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.01	114	899373	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.90	117	739078	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.44	65	324978	29.74	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.13%
60) 4-Bromofluorobenzene	11.92	95	286239	26.01	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	86.70%
63) Toluene-d8	9.56	98	1085021	31.80	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	106.00%

Target Compounds

					Qvalue
2) Dichlorodifluoromethane	1.82	85	39351	4.303 ug/l	95
3) Chloromethane	2.00	50	50088	4.728 ug/l	99
4) Vinyl Chloride	2.11	62	51621	4.790 ug/l	95
5) Bromomethane	2.43	94	39532	5.149 ug/l	96
6) Chloroethane	2.55	64	35086	5.021 ug/l	95
7) Trichlorofluoromethane	2.82	101	79914	4.973 ug/l	99
8) Diethyl Ether	3.14	74	30117	5.029 ug/l	98
9) 1,1,2-Trichlorotrifluoroet	3.44	101	51913	5.341 ug/l	87
10) 1,1-Dichloroethene	3.43	96	48470	5.043 ug/l	95
11) Methyl Iodide	3.63	142	64145	4.277 ug/l	95
12) Methyl Acetate	3.90	43	56830	5.677 ug/l	99
13) Acrolein	3.31	56	27819	25.197 ug/l	99
14) Acrylonitrile	4.45	53	100344	24.024 ug/l #	87
15) Acetone	3.48	58	31629	22.817 ug/l	91
16) Carbon Disulfide	3.72	76	126769	4.889 ug/l	96
17) Allyl chloride	3.91	41	74119	5.231 ug/l	98
18) Methylene Chloride	4.10	84	56673	5.201 ug/l	94
19) trans-1,2-Dichloroethene	4.51	96	54049	5.246 ug/l	92
20) Diisopropyl ether	5.32	45	148793	5.149 ug/l	93
21) 1,1-Dichloroethane	5.24	63	94653	5.259 ug/l	100
22) cis-1,2-Dichloroethene	6.21	96	60155	5.130 ug/l #	91
23) tert-Butyl Alcohol	4.24	59	32247	23.396 ug/l #	100
24) Methyl tert-Butyl Ether	4.49	73	138931	4.916 ug/l	97
25) Chloroform	6.76	83	99720	5.389 ug/l	97
26) Cyclohexane	7.07	56	77878	5.062 ug/l #	99
29) 1,1-Dichloropropene	7.20	75	68283	5.049 ug/l	97
30) 2-Butanone	6.20	43	128888	24.205 ug/l #	94
31) 2,2-Dichloropropane	6.21	77	74391	5.144 ug/l	96
32) 1,1,1-Trichloroethane	6.98	97	81396	5.197 ug/l	99
33) Carbon Tetrachloride	7.18	117	70624	5.065 ug/l	95
34) Benzene	7.46	78	222942	5.400 ug/l #	86
35) Methacrylonitrile	6.55	41	24512	4.555 ug/l	96
36) 1,2-Dichloroethane	7.54	62	70241	5.326 ug/l	99
37) Trichloroethene	8.27	130	63011	5.371 ug/l	100
38) Methylcyclohexane	8.52	83	80707	5.001 ug/l	97
39) 1,2-Dichloropropane	8.56	63	56964	3.118 ug/l	98
40) Dibromomethane	8.65	93	35134	5.128 ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081019\
 Data File : VN057164.D
 Acq On : 10 Aug 2019 9:50
 Operator : JC/SP
 Sample : VSTDIC005
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005

Quant Time: Aug 11 03:35:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N081019W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sun Aug 11 03:32:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Bromodichloromethane	8.85	83	66120	4.854	ug/l #	91
42) Vinyl Acetate	5.27	43	544297	24.588	ug/l	100
43) Ethyl Acetate	6.29	43	40551	4.129	ug/l	94
44) Isopropyl Acetate	7.56	43	90406	5.070	ug/l	97
45) 1,4-Dioxane	8.64	88	16119	96.871	ug/l	95
46) Methyl methacrylate	8.63	41	38302	4.711	ug/l	98
47) n-amyl Acetate	11.58	43	49423	3.767	ug/l	98
48) t-1,3-Dichloropropene	9.86	75	67294	4.633	ug/l	99
49) cis-1,3-Dichloropropene	9.30	75	78414	5.254	ug/l	96
50) 1,1,2-Trichloroethane	10.04	97	50350	5.142	ug/l	97
51) Ethyl methacrylate	9.90	69	62522	4.525	ug/l	99
52) 1,3-Dichloropropane	10.19	76	85156	5.218	ug/l	100
53) Dibromochloromethane	10.39	129	51522	4.670	ug/l	95
54) 1,2-Dibromoethane	10.50	107	49948	4.887	ug/l	99
55) 2-Chloroethyl vinyl ether	9.15	63	78421	20.100	ug/l	99
56) Bromoform	11.63	173	35782	4.389	ug/l	95
58) 4-Methyl-2-Pentanone	9.45	43	239509	25.226	ug/l	99
59) 2-Hexanone	10.24	43	166277	24.359	ug/l	97
61) Tetrachloroethene	10.11	164	72076	6.354	ug/l	97
62) Toluene	9.62	91	233540	5.551	ug/l	99
64) Chlorobenzene	10.93	112	135236	5.227	ug/l	97
65) 1,1,1,2-Tetrachloroethane	11.01	131	53584	5.400	ug/l	100
66) Ethyl Benzene	11.01	91	229345	5.086	ug/l	100
67) m/p-Xylenes	11.12	106	167783	9.866	ug/l	98
68) o-Xylene	11.46	106	81770	4.953	ug/l	94
69) Styrene	11.47	104	114489	4.325	ug/l	97
70) Isopropylbenzene	11.76	105	221338	4.979	ug/l	99
71) 1,1,2,2-Tetrachloroethane	12.02	83	64270	5.279	ug/l	95
72) 1,2,3-Trichloropropane	12.08	75	71628	6.636	ug/l	63
73) Bromobenzene	12.04	156	54836	4.757	ug/l	97
74) n-propylbenzene	12.11	91	202875	4.348	ug/l	98
75) 2-Chlorotoluene	12.20	91	137446	4.762	ug/l	98
76) 1,3,5-Trimethylbenzene	12.26	105	179751	5.002	ug/l	87
77) t-1,4-Dichloro-2-butene	11.82	75	14256	3.976	ug/l	88
78) 4-Chlorotoluene	12.30	91	112781	4.207	ug/l	99
79) tert-butylbenzene	12.52	119	168456	5.107	ug/l	98
80) 1,2,4-Trimethylbenzene	12.57	105	158566	4.492	ug/l	98
81) sec-Butylbenzene	12.70	105	192404	4.626	ug/l	99
82) p-Isopropyltoluene	12.82	119	163652	4.419	ug/l	100
83) 1,3-Dichlorobenzene	12.81	146	68398	3.897	ug/l	98
84) 1,4-Dichlorobenzene	12.89	146	58099	3.559	ug/l	99
85) n-Butylbenzene	13.15	91	96342	3.598	ug/l	99
86) Hexachloroethane	13.41	117	32526	4.830	ug/l	99
87) 1,2-Dichlorobenzene	13.19	146	69046	3.967	ug/l	98
88) 1,2-Dibromo-3-Chloropropan	13.80	75	3082	1.804	ug/l #	1
89) 1,2,4-Trichlorobenzene	14.45	180	14814	2.124	ug/l	95
90) Hexachlorobutadiene	14.55	225	39018	5.200	ug/l	95
91) Naphthalene	14.67	128	30987	2.345	ug/l	97
92) 1,2,3-Trichlorobenzene	14.84	180	13694	2.137	ug/l	88


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081019\
Data File : VN057164.D
Acq On : 10 Aug 2019 9:50
Operator : JC/SP
Sample : VSTDICC005
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Quant Time: Aug 11 03:35:41 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N081019W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sun Aug 11 03:32:59 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
 Bromochloromethane
 Concen: 30.000 ug/l
 RT: 6.58 min Scan# 1496
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Instrument :
 MSVOA_N
 ClientSampled :
 VSTDIC005

Tgt Ion:128 Resp: 167661
 Ion Ratio Lower Upper
 128 100
 49 141.7 0.0 361.0
 130 130.1 0.0 329.5




Abundance Ion 128.00 (127.70 to 128.70): V
 Ion 49.00 (48.70 to 49.70): VN
 Ion 130.00 (129.70 to 130.70): V






#2
 Dichlorodifluoromethane
 Concen: 4.303 ug/l
 RT: 1.82 min Scan# 14
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion: 85 Resp: 39351
 Ion Ratio Lower Upper
 85 100
 87 30.2 16.5 49.5



Abundance Ion 85.00 (84.70 to 85.70): VN
 Ion 87.00 (86.70 to 87.70): VN





#3

Chloromethane

Concen: 4.728 ug/l

RT: 2.00 min Scan# 69

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 50 Resp: 50088

Ion Ratio Lower Upper

50 100

52 32.6 26.5 39.7



#4

Vinyl Chloride

Concen: 4.790 ug/l

RT: 2.11 min Scan# 105

Delta R.T. -0.00 min

Lab File: VN057164.D




Acq: 10 Aug 2019 9:50


Tgt Ion: 62 Resp: 51621

Ion Ratio Lower Upper

62 100

64 33.9 25.0 37.4





#5

Bromomethane

Concen: 5.149 ug/l

RT: 2.43 min Scan# 205

Delta R.T. -0.01 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 94 Resp: 39532

Ion Ratio Lower Upper

94 100

96 97.1 74.7 112.1



#6

Chloroethane

Concen: 5.021 ug/l

RT: 2.55 min Scan# 243

Delta R.T. -0.00 min

Lab File: VN057164.D




Acq: 10 Aug 2019 9:50


Tgt Ion: 64 Resp: 35086

Ion Ratio Lower Upper

64 100

66 29.2 25.8 38.6






#7




Trichlorofluoromethane
Concen: 4.973 ug/l
RT: 2.82 min Scan# 327
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion: 101 Resp: 79914
Ion Ratio Lower Upper
101 100
103 64.1 51.8 77.8




Abundance Ion 101.00 (100.70 to 101.70): V
Ion 103.00 (102.70 to 103.70): V





#8


Diethyl Ether
Concen: 5.029 ug/l
RT: 3.14 min Scan# 425
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion: 74 Resp: 30117
Ion Ratio Lower Upper
74 100
45 91.5 17.9 161.5



Abundance Ion 74.00 (73.70 to 74.70): VNO
Ion 45.00 (44.70 to 45.70): VNO





#9

1,1,2-Trichlorotrifluoroethane

Concen: 5.341 ug/l

RT: 3.44 min Scan# 518

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005


Tgt Ion: 101 Resp: 51913

Ion Ratio Lower Upper

101 100

85 21.3 0.0 84.4




151 84.5 0.0 175.0



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VN0

Ion 151.00 (150.70 to 151.70): V



#10

1,1-Dichloroethene

Concen: 5.043 ug/l

RT: 3.43 min Scan# 516

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50


Tgt Ion: 96 Resp: 48470

Ion Ratio Lower Upper

96 100

61 143.0 118.2 177.2



98 67.0 49.3 73.9




Abundance Ion 96.00 (95.70 to 96.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 98.00 (97.70 to 98.70): VN0









#11
Methyl Iodide
Concen: 4.277 ug/l
RT: 3.63 min Scan# 576
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 142 Resp: 64145
Ion Ratio Lower Upper
142 100
127 40.6 19.0 57.0
141 12.3 7.8 23.5



#12
Methyl Acetate
Concen: 5.677 ug/l
RT: 3.90 min Scan# 660
Delta R.T. 0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion: 43 Resp: 56830
Ion Ratio Lower Upper
43 100
74 26.1 20.5 30.7





#13

Acrolein

Concen: 25.197 ug/l

RT: 3.31 min Scan# 479

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 56 Resp: 27819

Ion Ratio Lower Upper

56 100

55 69.8 56.5 84.7



#14

Acrylonitrile

Concen: 24.024 ug/l

RT: 4.45 min Scan# 833

Delta R.T. 0.01 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50




Tgt Ion: 53 Resp: 100344


Ion Ratio Lower Upper

53 100

52 78.8 41.1 123.4

51 15.7 17.4 52.3#





#15

Acetone

Concen: 22.817 ug/l

RT: 3.48 min Scan# 530

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :


VSTDIC005

Tgt Ion: 58 Resp: 31629

Ion Ratio Lower Upper


58 100

43 299.7 226.2 339.2





Abundance Ion 58.00 (57.70 to 58.70): VN057165.D (-516) (-)

Ion 43.00 (42.70 to 43.70): VN057165.D (-516) (-)



Time-->



#16

Carbon Disulfide

Concen: 4.889 ug/l

RT: 3.72 min Scan# 605

Delta R.T. -0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 76 Resp: 126769

Ion Ratio Lower Upper


76 100

78 7.5 7.3 10.9





Abundance Ion 76.00 (75.70 to 76.70): VN057165.D (-587) (-)

Ion 78.00 (77.70 to 78.70): VN057165.D (-587) (-)



Time-->









#17
 Allyl chloride
 Concen: 5.231 ug/l
 RT: 3.91 min Scan# 665
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 41 Resp: 74119
 Ion Ratio Lower Upper
 41 100
 39 64.3 31.2 93.6
 76 40.4 20.5 61.6



#18
 Methylene Chloride
 Concen: 5.201 ug/l
 RT: 4.10 min Scan# 723
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion: 84 Resp: 56673
 Ion Ratio Lower Upper
 84 100
 49 119.2 90.1 135.1
 51 40.1 27.9 41.9
 86 69.3 53.0 79.6





#19

trans-1,2-Dichloroethene

Concen: 5.246 ug/l

RT: 4.51 min Scan# 851

Delta R.T. 0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 96 Resp: 54049

Ion Ratio Lower Upper

96 100

61 121.4 103.6 155.4

98 73.1 51.3 76.9



#20

Diisopropyl ether

Concen: 5.149 ug/l

RT: 5.32 min Scan# 1102

Delta R.T. 0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Tgt Ion: 45 Resp: 148793




Ion Ratio Lower Upper


45 100

43 83.3 60.0 90.0

87 33.9 25.9 38.9

59 13.8 10.5 15.7









#21
 1,1-Dichloroethane
 Concen: 5.259 ug/l
 RT: 5.24 min Scan# 1079
 Delta R.T. -0.01 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion	Ratio	Lower	Upper
63	100		
98	7.0	3.5	10.4
100	3.6	1.9	5.7



#22
 cis-1,2-Dichloroethene
 Concen: 5.130 ug/l
 RT: 6.21 min Scan# 1380
 Delta R.T. 0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion	Ratio	Lower	Upper
96	100		
61	136.5	109.2	163.8
98	43.4	51.7	77.5#





#23

tert-Butyl Alcohol

Concen: 23.396 ug/l

RT: 4.24 min Scan# 767

Delta R.T. 0.01 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 59 Resp: 32247

Ion Ratio Lower Upper

59 100

52 0.0 0.2 0.2#



#24

Methyl tert-Butyl Ether

Concen: 4.916 ug/l

RT: 4.49 min Scan# 845

Delta R.T. -0.00 min

Lab File: VN057164.D




Acq: 10 Aug 2019 9:50


Tgt Ion: 73 Resp: 138931

Ion Ratio Lower Upper

73 100

43 19.2 14.3 21.5









#25
Chloroform
Concen: 5.389 ug/l
RT: 6.76 min Scan# 1550
Delta R.T. 0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 83 Resp: 99720
Ion Ratio Lower Upper
83 100
85 67.6 52.5 78.7



#26
Cyclohexane
Concen: 5.062 ug/l
RT: 7.07 min Scan# 1646
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion: 56 Resp: 77878
Ion Ratio Lower Upper
56 100
89 0.0 0.0 0.0
84 87.4 70.8 106.2





#27

1,2-Dichloroethane-d4

Concen: 29.735 ug/l

RT: 7.44 min Scan# 1761

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 65 Resp: 324978

Ion Ratio Lower Upper

65 100

67 55.0 42.4 63.6



#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 8.01 min Scan# 1939

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50




Tgt Ion: 114 Resp: 899373


Ion Ratio Lower Upper

114 100

63 19.2 14.6 21.8

88 15.8 12.2 18.2









#29
 1,1-Dichloropropene
 Concen: 5.049 ug/l
 RT: 7.20 min Scan# 1687
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion	Ratio	Lower	Upper
75	100		
110	40.2	0.0	75.6
77	32.2	0.0	62.4



#30
 2-Butanone
 Concen: 24.205 ug/l
 RT: 6.20 min Scan# 1378
 Delta R.T. 0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion	Ratio	Lower	Upper
43	100		
57	0.0	6.8	10.2#
72	28.7	23.0	34.6









#31
 2,2-Dichloropropane
 Concen: 5.144 ug/l
 RT: 6.21 min Scan# 1380
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 77 Resp: 74391
 Ion Ratio Lower Upper
 77 100
 97 26.1 19.4 29.2



#32
 1,1,1-Trichloroethane
 Concen: 5.197 ug/l
 RT: 6.98 min Scan# 1618
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion: 97 Resp: 81396
 Ion Ratio Lower Upper
 97 100
 99 65.6 52.4 78.6
 61 42.1 32.3 48.5





#33

Carbon Tetrachloride

Concen: 5.065 ug/l

RT: 7.18 min Scan# 1683

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005


Tgt Ion:117 Resp: 70624

Ion Ratio Lower Upper

117 100

119 97.1 73.9 110.9




121 32.4 23.4 35.2



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 119.00 (118.70 to 119.70): V

Ion 121.00 (120.70 to 121.70): V



#34

Benzene

Concen: 5.400 ug/l

RT: 7.46 min Scan# 1768

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50


Tgt Ion: 78 Resp: 222942

Ion Ratio Lower Upper

78 100

77 23.0 18.3 27.5



51 33.5 16.2 24.2#




Abundance Ion 78.10 (77.80 to 78.80): VNO

Ion 77.00 (76.70 to 77.70): VNO

Ion 51.00 (50.70 to 51.70): VNO









#35
Methacrylonitrile
Concen: 4.555 ug/l
RT: 6.55 min Scan# 1486
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:	41	Resp:	24512
Ion	Ratio	Lower	Upper
41	100		
39	45.1	24.6	74.0
67	66.5	34.4	103.2
52	32.5	14.9	44.5



#36
1,2-Dichloroethane
Concen: 5.326 ug/l
RT: 7.54 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion:	62	Resp:	70241
Ion	Ratio	Lower	Upper
62	100		
98	9.8	8.2	12.4









#37
 Trichloroethene
 Concen: 5.371 ug/l
 RT: 8.27 min Scan# 2020
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 130 Resp: 63011
 Ion Ratio Lower Upper
 130 100
 95 93.6 74.6 112.0



#38
 Methylcyclohexane
 Concen: 5.001 ug/l
 RT: 8.52 min Scan# 2099
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion: 83 Resp: 80707
 Ion Ratio Lower Upper
 83 100
 55 74.9 36.5 109.7
 98 49.0 23.4 70.0





#39

1,2-Dichloropropane

Concen: 3.118 ug/l

RT: 8.56 min Scan# 2110

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion: 63 Resp: 56964

Ion Ratio Lower Upper

63 100

65 33.9 26.2 39.4



#40

Dibromomethane

Concen: 5.128 ug/l

RT: 8.65 min Scan# 2140

Delta R.T. 0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50




Tgt Ion: 93 Resp: 35134


Ion Ratio Lower Upper

93 100

95 88.9 0.0 165.0

174 122.4 0.0 245.4





#41

Bromodichloromethane

Concen: 4.854 ug/l

RT: 8.85 min Scan# 2200

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005


Tgt Ion: 83 Resp: 66120

Ion Ratio Lower Upper

83 100

85 70.3 50.6 76.0




127 11.0 7.0 10.6#



Abundance Ion 83.00 (82.70 to 83.70): VN057164.D (-2185) (-)

Ion 85.00 (84.70 to 85.70): VN057164.D (-2185) (-)

Ion 127.00 (126.70 to 127.70): VN057164.D (-2185) (-)



#42

Vinyl Acetate

Concen: 24.588 ug/l

RT: 5.27 min Scan# 1086

Delta R.T. 0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 43 Resp: 544297

Ion Ratio Lower Upper



43 100


86 9.8 8.0 12.0



Abundance Ion 43.00 (42.70 to 43.70): VN057164.D (-930) (-)

Ion 86.00 (85.70 to 86.70): VN057164.D (-930) (-)









#43
Ethyl Acetate
Concen: 4.129 ug/l
RT: 6.29 min Scan# 1404
Delta R.T. 0.01 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 40551
Ion Ratio Lower Upper
43 100
45 10.1 10.1 15.1



#44
Isopropyl Acetate
Concen: 5.070 ug/l
RT: 7.56 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion: 43 Resp: 90406
Ion Ratio Lower Upper
43 100
61 29.2 22.2 33.4









#45
 1,4-Dioxane
 Concen: 96.871 ug/l
 RT: 8.64 min Scan# 2137
 Delta R.T. 0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 88 Resp: 16119
 Ion Ratio Lower Upper
 88 100
 43 23.6 20.5 30.7
 58 71.3 53.8 80.8



#46
 Methyl methacrylate
 Concen: 4.711 ug/l
 RT: 8.63 min Scan# 2134
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion: 41 Resp: 38302
 Ion Ratio Lower Upper
 41 100
 69 93.2 48.1 144.4
 39 53.4 26.5 79.5





#47

n-amyl Acetate

Concen: 3.767 ug/l

RT: 11.58 min Scan# 3051

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :


VSTDIC005

Tgt Ion: 43 Resp: 49423

Ion Ratio Lower Upper




43 100

55 26.8 20.7 31.1



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 55.00 (54.70 to 55.70): VN0



#48

t-1,3-Dichloropropene

Concen: 4.633 ug/l

RT: 9.86 min Scan# 2514

Delta R.T. 0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 75 Resp: 67294

Ion Ratio Lower Upper



75 100


77 31.8 24.9 37.3



Abundance Ion 75.00 (74.70 to 75.70): VN0

Ion 77.00 (76.70 to 77.70): VN0









#49
 cis-1,3-Dichloropropene
 Concen: 5.254 ug/l
 RT: 9.30 min Scan# 2340
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 75 Resp: 78414
 Ion Ratio Lower Upper
 75 100
 77 31.4 25.8 38.8
 39 44.7 33.0 49.6



#50
 1,1,2-Trichloroethane
 Concen: 5.142 ug/l
 RT: 10.04 min Scan# 2571
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion: 97 Resp: 50350
 Ion Ratio Lower Upper
 97 100
 83 89.4 71.4 107.0
 85 54.6 43.8 65.8
 99 71.2 51.0 76.6





#51

Ethyl methacrylate

Concen: 4.525 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005


Tgt Ion: 69 Resp: 62522

Ion Ratio Lower Upper

69 100

41 61.1 30.4 91.2




39 30.1 14.8 44.4



Abundance Ion 69.00 (68.70 to 69.70): VN057165.D (-2514) (-)

Ion 41.00 (40.70 to 41.70): VN057165.D (-2514) (-)

Ion 39.00 (38.70 to 39.70): VN057165.D (-2514) (-)



#52

1,3-Dichloropropane

Concen: 5.218 ug/l

RT: 10.19 min Scan# 2618

Delta R.T. 0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 76 Resp: 85156

Ion Ratio Lower Upper



76 100


78 32.5 0.0 65.0



Abundance Ion 76.00 (75.70 to 76.70): VN057165.D (-2604) (-)

Ion 78.00 (77.70 to 78.70): VN057165.D (-2604) (-)









#53
 Dibromochloromethane
 Concen: 4.670 ug/l
 RT: 10.39 min Scan# 2679
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:129 Resp: 51522
 Ion Ratio Lower Upper
 129 100
 127 80.8 38.2 114.6



#54
 1,2-Dibromoethane
 Concen: 4.887 ug/l
 RT: 10.50 min Scan# 2713
 Delta R.T. 0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion:107 Resp: 49948
 Ion Ratio Lower Upper
 107 100
 109 97.0 76.8 115.2









#55
 2-Chloroethyl vinyl ether
 Concen: 20.100 ug/l
 RT: 9.15 min Scan# 2293
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampled :
 VSTDIC005


Tgt Ion: 63 Resp: 78421
 Ion Ratio Lower Upper
 63 100
 65 32.6 26.0 39.0
 106 27.9 23.3 34.9



#56
 Bromoform
 Concen: 4.389 ug/l
 RT: 11.63 min Scan# 3067
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion: 173 Resp: 35782
 Ion Ratio Lower Upper
 173 100
 175 52.8 39.0 58.4
 254 10.1 8.6 13.0









#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.90 min Scan# 2840
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50




Instrument :
MSVOA_N
ClientSampled :
VSTDIC005


Tgt Ion: 117 Resp: 739078
Ion Ratio Lower Upper
117 100
82 51.5 41.5 62.3
119 32.1 26.1 39.1



#58
4-Methyl-2-Pentanone
Concen: 25.226 ug/l
RT: 9.45 min Scan# 2386
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion: 43 Resp: 239509
Ion Ratio Lower Upper
43 100
58 41.6 32.8 49.2
85 20.2 16.4 24.6









#59
2-Hexanone
Concen: 24.359 ug/l
RT: 10.24 min Scan# 2632
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 166277
Ion Ratio Lower Upper
43 100
58 54.8 45.4 68.2
57 18.8 16.1 24.1



#60
4-Bromofluorobenzene
Concen: 26.010 ug/l
RT: 11.92 min Scan# 3155
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion: 95 Resp: 286239
Ion Ratio Lower Upper
95 100
174 93.4 74.9 112.3
176 90.6 72.1 108.1





#61

Tetrachloroethene

Concen: 6.354 ug/l

RT: 10.11 min Scan# 2593

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:164 Resp: 72076


Ion Ratio Lower Upper

164 100

166 126.0 103.0 154.6

129 89.8 67.3 100.9

131 87.0 68.5 102.7






Abundance

Ion 163.85 (163.55 to 164.55): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.90 (128.60 to 129.60): V

Ion 130.90 (130.60 to 131.60): V



#62

Toluene

Concen: 5.551 ug/l

RT: 9.62 min Scan# 2441

Delta R.T. -0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 91 Resp: 233540

Ion Ratio Lower Upper

91 100



92 58.3 46.1 69.1




Abundance

Ion 91.00 (90.70 to 91.70): VNO

Ion 92.00 (91.70 to 92.70): VNO





#63

Toluene-d8

Concen: 31.802 ug/l

RT: 9.56 min Scan# 2421

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :


VSTDIC005

Tgt Ion: 98 Resp: 1085021

Ion Ratio Lower Upper

98 100

100 64.1 52.4 78.6



Abundance Ion 98.00 (97.70 to 98.70): VN057164.D (-2406) (-)

Ion 100.00 (99.70 to 100.70): VN057164.D (-2406) (-)

9.56

600000



400000

200000

0

9.50 9.55 9.60 9.70

Time-->



#64

Chlorobenzene

Concen: 5.227 ug/l

RT: 10.93 min Scan# 2848

Delta R.T. -0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 112 Resp: 135236

Ion Ratio Lower Upper

112 100

114 33.2 25.0 37.6



Abundance Ion 112.00 (111.70 to 112.70): VN057164.D (-2834) (-)

Ion 114.00 (113.70 to 114.70): VN057164.D (-2834) (-)

10.93

80000

60000


40000


20000

0

10.85 10.90 10.95 11.00

Time-->





#65

1,1,1,2-Tetrachloroethane

Concen: 5.400 ug/l

RT: 11.01 min Scan# 2873

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005


Tgt Ion:131 Resp: 53584

Ion Ratio Lower Upper

131 100

133 97.9 0.0 196.0

119 65.5 0.0 132.6






Abundance

Ion 131.00 (130.70 to 131.70): V

Ion 133.00 (132.70 to 133.70): V

Ion 119.00 (118.70 to 119.70): V



#66

Ethyl Benzene

Concen: 5.086 ug/l

RT: 11.01 min Scan# 2873

Delta R.T. -0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 91 Resp: 229345

Ion Ratio Lower Upper

91 100



106 31.8 25.3 37.9




Abundance

Ion 91.00 (90.70 to 91.70): V

Ion 106.00 (105.70 to 106.70): V









#67
m/p-Xylenes
Concen: 9.866 ug/l
RT: 11.12 min Scan# 2908
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:106 Resp: 167783
Ion Ratio Lower Upper
106 100
91 198.2 160.5 240.7



#68
o-Xylene
Concen: 4.953 ug/l
RT: 11.46 min Scan# 3012
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion:106 Resp: 81770
Ion Ratio Lower Upper
106 100
91 220.2 105.3 315.8





#69

Styrene

Concen: 4.325 ug/l

RT: 11.47 min Scan# 3017

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:104 Resp: 114489

Ion Ratio Lower Upper

104 100

78 49.3 38.7 58.1

103 58.7 44.3 66.5



#70

Isopropylbenzene

Concen: 4.979 ug/l

RT: 11.76 min Scan# 3107

Delta R.T. -0.00 min

Lab File: VN057164.D




Acq: 10 Aug 2019 9:50


Tgt Ion:105 Resp: 221338

Ion Ratio Lower Upper

105 100

120 27.7 19.1 35.5









#71
 1,1,2,2-Tetrachloroethane
 Concen: 5.279 ug/l
 RT: 12.02 min Scan# 3188
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 83 Resp: 64270
 Ion Ratio Lower Upper
 83 100
 131 9.9 5.5 16.5
 85 61.5 32.6 98.0



#72
 1,2,3-Trichloropropane
 Concen: 6.636 ug/l
 RT: 12.08 min Scan# 3204
 Delta R.T. 0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion: 75 Resp: 71628
 Ion Ratio Lower Upper
 75 100
 77 135.0 0.0 379.4





#73

Bromobenzene

Concen: 4.757 ug/l

RT: 12.04 min Scan# 3194

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005


Tgt Ion:156 Resp: 54836

Ion Ratio Lower Upper

156 100

77 176.3 0.0 361.6




158 97.5 0.0 200.6



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VN0

Ion 158.00 (157.70 to 158.70): V



#74

n-propylbenzene

Concen: 4.348 ug/l

RT: 12.11 min Scan# 3216

Delta R.T. -0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 91 Resp: 202875

Ion Ratio Lower Upper



91 100


120 23.4 0.0 48.2



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 120.00 (119.70 to 120.70): V





#75

2-Chlorotoluene

Concen: 4.762 ug/l

RT: 12.20 min Scan# 3242

Delta R.T. 0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :


VSTDIC005

Tgt Ion: 91 Resp: 137446

Ion Ratio Lower Upper




91 100

126 37.6 0.0 73.4



Abundance Ion 91.00 (90.70 to 91.70): VN057165.D (-3229) (-)

Ion 126.00 (125.70 to 126.70): VN057165.D (-3229) (-)



#76

1,3,5-Trimethylbenzene

Concen: 5.002 ug/l

RT: 12.26 min Scan# 3260

Delta R.T. -0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 105 Resp: 179751

Ion Ratio Lower Upper



105 100


120 50.1 0.0 83.6



Abundance Ion 105.00 (104.70 to 105.70): VN057165.D (-3246) (-)

Ion 120.00 (119.70 to 120.70): VN057165.D (-3246) (-)





#77

t-1,4-Dichloro-2-butene

Concen: 3.976 ug/l

RT: 11.82 min Scan# 3124

Delta R.T. 0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005


Tgt Ion: 75 Resp: 14256

Ion Ratio Lower Upper

75 100

53 106.1 45.4 136.2




89 50.0 23.5 70.6



Abundance Ion 75.00 (74.70 to 75.70): VN057164.D (-3118) (-)

Ion 53.00 (52.70 to 53.70): VN057164.D (-3118) (-)

Ion 89.00 (88.70 to 89.70): VN057164.D (-3118) (-)



#78

4-Chlorotoluene

Concen: 4.207 ug/l

RT: 12.30 min Scan# 3273

Delta R.T. -0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 91 Resp: 112781

Ion Ratio Lower Upper



91 100


126 34.7 0.0 70.2



Abundance Ion 91.00 (90.70 to 91.70): VN057164.D (-3118) (-)

Ion 126.00 (125.70 to 126.70): VN057164.D (-3118) (-)









#79
 tert-butylbenzene
 Concen: 5.107 ug/l
 RT: 12.52 min Scan# 3342
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Instrument :
 MSVOA_N
 ClientSampled :
 VSTDIC005

Tgt Ion:119 Resp: 168456
 Ion Ratio Lower Upper
 119 100
 91 60.6 0.0 123.8
 134 24.9 0.0 48.6




Abundance Ion 119.00 (118.70 to 119.70): V
 Ion 91.00 (90.70 to 91.70): V
 Ion 134.00 (133.70 to 134.70): V






#80
 1,2,4-Trimethylbenzene
 Concen: 4.492 ug/l
 RT: 12.57 min Scan# 3357
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion:105 Resp: 158566
 Ion Ratio Lower Upper
 105 100
 120 46.7 0.0 91.2



Abundance Ion 105.00 (104.70 to 105.70): V
 Ion 120.00 (119.70 to 120.70): V









#81
 sec-Butylbenzene
 Concen: 4.626 ug/l
 RT: 12.70 min Scan# 3398
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 192404
 Ion Ratio Lower Upper
 105 100
 134 20.9 0.0 40.6



#82
 p-Isopropyltoluene
 Concen: 4.419 ug/l
 RT: 12.82 min Scan# 3436
 Delta R.T. 0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion:119 Resp: 163652
 Ion Ratio Lower Upper
 119 100
 134 27.0 0.0 54.2
 91 21.9 0.0 44.4





#83

1,3-Dichlorobenzene

Concen: 3.897 ug/l

RT: 12.81 min Scan# 3433

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampled :

VSTDIC005


Tgt Ion:146 Resp: 68398

Ion Ratio Lower Upper

146 100

111 36.0 19.1 57.1




148 64.4 32.6 97.8



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#84

1,4-Dichlorobenzene

Concen: 3.559 ug/l

RT: 12.89 min Scan# 3458

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50


Tgt Ion:146 Resp: 58099

Ion Ratio Lower Upper

146 100

111 36.4 18.0 54.0



148 63.5 32.3 96.9




Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V





#85

n-Butylbenzene

Concen: 3.598 ug/l

RT: 13.15 min Scan# 3537

Delta R.T. -0.00 min

Lab File: VN057164.D

Acq: 10 Aug 2019 9:50

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005


Tgt Ion: 91 Resp: 96342

Ion Ratio Lower Upper

91 100

92 53.3 0.0 104.4




134 28.1 0.0 55.0



Abundance Ion 91.00 (90.70 to 91.70): VN057165.D (-3524) (-)

Ion 92.00 (91.70 to 92.70): VN057165.D (-3524) (-)

Ion 134.00 (133.70 to 134.70): VN057165.D (-3524) (-)



#86

Hexachloroethane

Concen: 4.830 ug/l

RT: 13.41 min Scan# 3619

Delta R.T. -0.00 min

Lab File: VN057164.D


Acq: 10 Aug 2019 9:50

Tgt Ion: 117 Resp: 32526

Ion Ratio Lower Upper



117 100


201 97.9 49.3 147.9



Abundance Ion 117.00 (116.70 to 117.70): VN057165.D (-3605) (-)

Ion 201.00 (200.70 to 201.70): VN057165.D (-3605) (-)









#87
 1,2-Dichlorobenzene
 Concen: 3.967 ug/l
 RT: 13.19 min Scan# 3550
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50




Instrument :
 MSVOA_N
 ClientSampled :
 VSTDIC005


Tgt Ion:	146	Resp:	69046
Ion	Ratio	Lower	Upper
146	100		
111	38.0	18.9	56.5
148	62.3	32.4	97.0



#88
 1,2-Dibromo-3-Chloropropane
 Concen: 1.804 ug/l
 RT: 13.80 min Scan# 3741
 Delta R.T. -0.01 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion:	75	Resp:	3082
Ion	Ratio	Lower	Upper
75	100		
155	198.2	80.9	121.3#
157	262.2	111.3	166.9#









#89
 1,2,4-Trichlorobenzene
 Concen: 2.124 ug/l
 RT: 14.45 min Scan# 3943
 Delta R.T. 0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Instrument :
 MSVOA_N
 ClientSampled :
 VSTDIC005

Tgt Ion:180 Resp: 14814
 Ion Ratio Lower Upper
 180 100
 182 89.5 75.0 112.6
 145 30.1 21.7 32.5




Abundance Ion 180.00 (179.70 to 180.70): V
 Ion 182.00 (181.70 to 182.70): V
 Ion 145.00 (144.70 to 145.70): V






#90
 Hexachlorobutadiene
 Concen: 5.200 ug/l
 RT: 14.55 min Scan# 3973
 Delta R.T. -0.00 min
 Lab File: VN057164.D
 Acq: 10 Aug 2019 9:50

Tgt Ion:225 Resp: 39018
 Ion Ratio Lower Upper
 225 100
 223 60.2 0.0 128.0
 227 59.5 0.0 125.6



Abundance Ion 225.00 (224.70 to 225.70): V
 Ion 223.00 (222.70 to 223.70): V
 Ion 227.00 (226.70 to 227.70): V









#91
Naphthalene
Concen: 2.345 ug/l
RT: 14.67 min Scan# 4010
Delta R.T. -0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion:128 Resp: 30987
Ion Ratio Lower Upper
128 100
127 10.9 10.2 15.4
129 10.9 8.6 13.0



#92
1,2,3-Trichlorobenzene
Concen: 2.137 ug/l
RT: 14.84 min Scan# 4063
Delta R.T. 0.00 min
Lab File: VN057164.D
Acq: 10 Aug 2019 9:50

Tgt Ion:180 Resp: 13694
Ion Ratio Lower Upper
180 100
182 109.4 0.0 195.8
145 38.5 0.0 61.2

