

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081020\  
 Data File : VN062807.D  
 Acq On : 10 Aug 2020 14:32  
 Operator : JC/MD  
 Sample : L3618-02  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BP-VPB179-GW-438-440

Quant Time: Aug 11 06:54:37 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N080520W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Aug 05 13:26:42 2020  
 Response via : Initial Calibration


| Internal Standards             | R.T.  | QIon | Response | Conc      | Units  | Dev(Min) |
|--------------------------------|-------|------|----------|-----------|--------|----------|
| 1) Pentafluorobenzene          | 7.63  | 168  | 156728   | 50.00     | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene        | 8.55  | 114  | 273812   | 50.00     | ug/l   | 0.00     |
| 63) Chlorobenzene-d5           | 11.38 | 117  | 272828   | 50.00     | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4     | 13.32 | 152  | 112376   | 50.00     | ug/l   | 0.00     |
| System Monitoring Compounds    |       |      |          |           |        |          |
| 33) 1,2-Dichloroethane-d4      | 7.99  | 65   | 122567   | 47.29     | ug/l   | 0.00     |
| Spiked Amount 50.000           |       |      | Recovery | =         | 94.58% |          |
| 35) Dibromofluoromethane       | 7.55  | 113  | 90889    | 48.53     | ug/l   | 0.00     |
| Spiked Amount 50.000           |       |      | Recovery | =         | 97.06% |          |
| 50) Toluene-d8                 | 10.06 | 98   | 351553   | 48.08     | ug/l   | 0.00     |
| Spiked Amount 50.000           |       |      | Recovery | =         | 96.16% |          |
| 62) 4-Bromofluorobenzene       | 12.38 | 95   | 134944   | 48.83     | ug/l   | 0.00     |
| Spiked Amount 50.000           |       |      | Recovery | =         | 97.66% |          |
| Target Compounds               |       |      |          |           |        |          |
|                                |       |      |          |           | Qvalue |          |
| 2) Dichlorodifluoromethane     | 1.85  | 85   | 476      | 0.212     | ug/l # | 42       |
| 3) Chloromethane               | 2.04  | 50   | 1891     | 0.893     | ug/l   | 93       |
| 11) Tert butyl alcohol         | 4.68  | 59   | 135      | 0.318     | ug/l # | 72       |
| 12) 1,1-Dichloroethene         | 3.68  | 96   | 356      | Below Cal | #      | 13       |
| 14) Allyl chloride             | 4.03  | 41   | 1749     | 0.637     | ug/l # | 75       |
| 16) Acetone                    | 3.78  | 43   | 11499    | 12.279    | ug/l # | 68       |
| 17) Carbon Disulfide           | 4.01  | 76   | 408      | Below Cal | #      | 74       |
| 18) Methyl Acetate             | 4.28  | 43   | 633      | Below Cal | #      | 83       |
| 20) Methylene Chloride         | 4.48  | 84   | 99       | Below Cal | #      | 5        |
| 25) 2-Butanone                 | 6.80  | 43   | 2902     | 2.188     | ug/l   | 99       |
| 31) Cyclohexane                | 7.62  | 56   | 3563     | 1.134     | ug/l # | 6        |
| 36) 1,1-Dichloropropene        | 7.62  | 75   | 13326    | 4.998     | ug/l # | 52       |
| 38) Carbon Tetrachloride       | 7.63  | 117  | 16428    | 5.210     | ug/l # | 15       |
| 44) Trichloroethene            | 8.80  | 130  | 15219    | 7.852     | ug/l   | 96       |
| 48) Methyl methacrylate        | 9.14  | 41   | 497      | 0.218     | ug/l # | 20       |
| 59) 2-Hexanone                 | 10.73 | 43   | 629      | 0.281     | ug/l # | 25       |
| 60) Dibromochloromethane       | 10.61 | 129  | 2007     | 0.838     | ug/l # | 10       |
| 64) Tetrachloroethene          | 10.61 | 164  | 1634     | 0.928     | ug/l # | 88       |
| 71) Bromoform                  | 12.38 | 173  | 531      | 0.283     | ug/l # | 1        |
| 76) 1,2,3-Trichloropropane     | 12.38 | 75   | 70683    | 23.323    | ug/l # | 100      |
| 81) trans-1,4-Dichloro-2-buten | 12.38 | 75   | 70683    | 71.198    | ug/l # | 10       |
| 95) Naphthalene                | 15.11 | 128  | 712      | 3.836     | ug/l # | 68       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081020\  
Data File : VN062807.D  
Acq On : 10 Aug 2020 14:32  
Operator : JC/MD  
Sample : L3618-02  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BP-VPB179-GW-438-440

Quant Time: Aug 11 06:54:37 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_N\METHODS\82N080520W.M  
Quant Title : SW846 8260  
QLast Update : Wed Aug 05 13:26:42 2020  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.00 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32

Instrument :

MSVOA\_N

ClientSampleId :


BP-VPB179-GW-438-440

Tot Ion:168 Resp: 156728

Ion Ratio Lower Upper




168 100

99 62.7 52.9 79.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#2

Dichlorodifluoromethane

Concen: 0.212 ug/l

RT: 1.85 min Scan# 25

Delta R.T. 0.02 min

Lab File: VN062807.D


Acq: 10 Aug 2020 14:32

Tgt Ion: 85 Resp: 476

Ion Ratio Lower Upper



85 100


87 0.0 16.2 48.6#



Abundance Ion 84.90 (84.60 to 85.60): VN0

Ion 86.90 (86.60 to 87.60): VN0





#3

Chloromethane

Concen: 0.893 ug/l

RT: 2.04 min Scan# 82

Delta R.T. 0.00 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32

Instrument :

MSVOA\_N

ClientSampleId :





BP-VPB179-GW-438-440

Tot Ion: 50 Resp: 1891

Ion Ratio Lower Upper

50 100

52 36.0 25.6 38.4



#11

Tert butyl alcohol

Concen: 0.318 ug/l

RT: 4.68 min Scan# 905

Delta R.T. -0.06 min

Lab File: VN062807.D




Acq: 10 Aug 2020 14:32


Tgt Ion: 59 Resp: 135

Ion Ratio Lower Upper

59 100

57 0.0 8.3 12.5#









#12  
 1,1-Dichloroethene  
 Concen: Below Cal  
 RT: 3.68 min Scan# 594  
 Delta R.T. -0.01 min  
 Lab File: VN062807.D  
 Acq: 10 Aug 2020 14:32




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BP-VPB179-GW-438-440


Tot Ion: 96 Resp: 356  
 Ion Ratio Lower Upper  
 96 100  
 61 50.7 139.0 208.6#  
 98 0.0 50.2 75.4#



#14  
 Allyl chloride  
 Concen: 0.637 ug/l  
 RT: 4.03 min Scan# 703  
 Delta R.T. -0.24 min  
 Lab File: VN062807.D  
 Acq: 10 Aug 2020 14:32

Tgt Ion: 41 Resp: 1749  
 Ion Ratio Lower Upper  
 41 100  
 39 62.9 55.7 83.5  
 76 0.0 27.5 41.3#









#16  
Acetone  
Concen: 12.279 ug/l  
RT: 3.78 min Scan# 624  
Delta R.T. 0.00 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32




Instrument :  
MSVOA\_N  
ClientSampleId :  
BP-VPB179-GW-438-440


Tot Ion: 43 Resp: 11499  
Ion Ratio Lower Upper  
43 100  
58 47.3 24.1 36.1#



#17  
Carbon Disulfide  
Concen: Below Cal  
RT: 4.01 min Scan# 695  
Delta R.T. 0.00 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32

Tgt Ion: 76 Resp: 408  
Ion Ratio Lower Upper  
76 100  
78 0.0 7.7 11.5#









#18  
Methvl Acetate  
Concen: Below Cal  
RT: 4.28 min Scan# 779  
Delta R.T. 0.00 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32




Instrument :  
MSVOA\_N  
ClientSampleId :  
BP-VPB179-GW-438-440


Tot Ion: 43 Resp: 633  
Ion Ratio Lower Upper  
43 100  
74 14.8 18.6 28.0#



#20  
Methylene Chloride  
Concen: Below Cal  
RT: 4.48 min Scan# 843  
Delta R.T. -0.01 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32

Tgt Ion: 84 Resp: 99  
Ion Ratio Lower Upper  
84 100  
49 0.0 103.4 155.2#  
51 0.0 29.3 43.9#  
86 0.0 49.9 74.9#









#25  
2-Butanone  
Concen: 2.188 ug/l  
RT: 6.80 min Scan# 1563  
Delta R.T. 0.01 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32




Instrument :  
MSVOA\_N  
ClientSampleId :  
BP-VPB179-GW-438-440


Tot Ion: 43 Resp: 2902  
Ion Ratio Lower Upper  
43 100  
72 25.8 21.1 31.7



#31  
Cyclohexane  
Concen: 1.134 ug/l  
RT: 7.62 min Scan# 1818  
Delta R.T. 0.01 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32

Tgt Ion: 56 Resp: 3563  
Ion Ratio Lower Upper  
56 100  
69 172.1 28.1 42.1#  
84 114.7 66.8 100.2#









#33  
 1,2-Dichloroethane-d4  
 Concen: 47.291 ug/l  
 RT: 7.99 min Scan# 1933  
 Delta R.T. -0.00 min  
 Lab File: VN062807.D  
 Acq: 10 Aug 2020 14:32




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BP-VPB179-GW-438-440


Tot Ion: 65 Resp: 122567  
 Ion Ratio Lower Upper  
 65 100  
 67 51.9 0.0 101.4



#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 8.55 min Scan# 2108  
 Delta R.T. -0.00 min  
 Lab File: VN062807.D  
 Acq: 10 Aug 2020 14:32

Tgt Ion: 114 Resp: 273812  
 Ion Ratio Lower Upper  
 114 100  
 63 22.2 0.0 47.6  
 88 17.9 0.0 34.4









#35  
 Dibromofluoromethane  
 Concen: 48.525 ug/l  
 RT: 7.55 min Scan# 1797  
 Delta R.T. -0.00 min  
 Lab File: VN062807.D  
 Acq: 10 Aug 2020 14:32

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BP-VPB179-GW-438-440

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 113     | 100   |       |       |
| 111     | 102.7 | 81.8  | 122.6 |
| 192     | 18.9  | 15.0  | 22.4  |




Abundance Ion 112.80 (112.50 to 113.50): V  
 Ion 110.80 (110.50 to 111.50): V  
 Ion 191.70 (191.40 to 192.40): V






#36  
 1,1-Dichloropropene  
 Concen: 4.998 ug/l  
 RT: 7.62 min Scan# 1819  
 Delta R.T. -0.13 min  
 Lab File: VN062807.D  
 Acq: 10 Aug 2020 14:32

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 75      | 100   |       |       |
| 110     | 9.0   | 15.6  | 46.7# |
| 77      | 0.0   | 24.1  | 36.1# |



Abundance Ion 74.90 (74.60 to 75.60): V  
 Ion 109.90 (109.60 to 110.60): V  
 Ion 76.90 (76.60 to 77.60): V





#38

Carbon Tetrachloride

Concen: 5.210 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.11 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32

Instrument :

MSVOA\_N

ClientSampleId :

BP-VPB179-GW-438-440


Tot Ion:117 Resp: 16428

Ion Ratio Lower Upper

117 100

119 6.0 79.4 119.2#




121 0.0 24.9 37.3#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#44

Trichloroethene

Concen: 7.852 ug/l

RT: 8.80 min Scan# 2186

Delta R.T. -0.00 min

Lab File: VN062807.D


Acq: 10 Aug 2020 14:32

Tgt Ion:130 Resp: 15219

Ion Ratio Lower Upper



130 100


95 108.2 0.0 207.6



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): V









#48  
Methvl methacrylate  
Concen: 0.218 ug/l  
RT: 9.14 min Scan# 2291  
Delta R.T. -0.03 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32




Instrument :  
MSVOA\_N  
ClientSampleId :  
BP-VPB179-GW-438-440


Tot Ion: 41 Resp: 497  
Ion Ratio Lower Upper  
41 100  
69 0.0 67.4 101.2#  
39 13.5 47.6 71.4#



#50  
Toluene-d8  
Concen: 48.076 ug/l  
RT: 10.06 min Scan# 2578  
Delta R.T. -0.00 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32

Tgt Ion: 98 Resp: 351553  
Ion Ratio Lower Upper  
98 100  
100 63.4 50.9 76.3





#59

2-Hexanone

Concen: 0.281 ug/l

RT: 10.73 min Scan# 2785

Delta R.T. 0.00 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32

Instrument :

MSVOA\_N

ClientSampleId :





BP-VPB179-GW-438-440

Tot Ion: 43 Resp: 629

Ion Ratio Lower Upper

43 100

58 0.0 26.7 80.1#



#60

Dibromochloromethane

Concen: 0.838 ug/l

RT: 10.61 min Scan# 2747

Delta R.T. -0.27 min

Lab File: VN062807.D




Acq: 10 Aug 2020 14:32


Tgt Ion: 129 Resp: 2007

Ion Ratio Lower Upper

129 100

127 0.0 39.1 117.2#





#62

4-Bromofluorobenzene

Concen: 48.833 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.00 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32

Instrument :

MSVOA\_N

ClientSampleId :

BP-VPB179-GW-438-440


Tot Ion: 95 Resp: 134944

Ion Ratio Lower Upper

95 100

174 75.6 0.0 144.8




176 71.0 0.0 142.2



Abundance Ion 94.90 (94.60 to 95.60): VN062807.D (-3205) (-)

Ion 173.80 (173.50 to 174.50): VN062807.D (-3205) (-)

Ion 175.80 (175.50 to 176.50): VN062807.D (-3205) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.38 min Scan# 2988

Delta R.T. -0.00 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32


Tgt Ion: 117 Resp: 272828

Ion Ratio Lower Upper

117 100

82 60.0 49.0 73.4



119 31.8 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): VN062807.D (-2895) (-)

Ion 82.00 (81.70 to 82.70): VN062807.D (-2895) (-)

Ion 118.90 (118.60 to 119.60): VN062807.D (-2895) (-)









#64  
Tetrachloroethene  
Concen: 0.928 ug/l  
RT: 10.61 min Scan# 2747  
Delta R.T. 0.00 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32




Instrument :  
MSVOA\_N  
ClientSampleId :  
BP-VPB179-GW-438-440


Tot Ion:164 Resp: 1634  
Ion Ratio Lower Upper  
164 100  
166 115.4 103.8 155.8  
129 99.3 82.7 124.1  
131 77.5 78.2 117.2#



#71  
Bromoform  
Concen: 0.283 ug/l  
RT: 12.38 min Scan# 3298  
Delta R.T. 0.27 min  
Lab File: VN062807.D  
Acq: 10 Aug 2020 14:32

Tgt Ion:173 Resp: 531  
Ion Ratio Lower Upper  
173 100  
175 1234.5 24.0 72.0#  
254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. -0.00 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32

Instrument :

MSVOA\_N

ClientSampleId :

BP-VPB179-GW-438-440


Tot Ion:152 Resp: 112376

Ion Ratio Lower Upper

152 100

115 62.3 31.6 94.8




150 155.3 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 23.323 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.15 min

Lab File: VN062807.D


Acq: 10 Aug 2020 14:32

Tgt Ion: 75 Resp: 70683

Ion Ratio Lower Upper



75 100


77 1.1 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0





#81

trans-1,4-Dichloro-2-butene

Concen: 71.198 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. 0.10 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32

Instrument :

MSVOA\_N

ClientSampleId :

BP-VPB179-GW-438-440


Tot Ion: 75 Resp: 70683

Ion Ratio Lower Upper

75 100

53 0.0 79.1 118.7#




89 0.0 36.2 54.2#



Abundance Ion 74.90 (74.60 to 75.60): VN062807.D (-3174) (-)

Ion 53.00 (52.70 to 53.70): VN062807.D (-3174) (-)

Ion 88.90 (88.60 to 89.60): VN062807.D (-3174) (-)



#95

Naphthalene

Concen: 3.836 ug/l

RT: 15.11 min Scan# 4148

Delta R.T. 0.01 min

Lab File: VN062807.D

Acq: 10 Aug 2020 14:32


Tgt Ion: 128 Resp: 712

Ion Ratio Lower Upper

128 100

127 0.0 10.8 16.2#

129 0.0 8.8 13.2#



Abundance Ion 128.00 (127.70 to 128.70): VN062807.D (-4053) (-)

Ion 127.00 (126.70 to 127.70): VN062807.D (-4053) (-)

Ion 129.00 (128.70 to 129.70): VN062807.D (-4053) (-)

