

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081020\
 Data File : VN062801.D
 Acq On : 10 Aug 2020 11:47
 Operator : JC/MD
 Sample : L3563-01MEDL 10X
 Misc : 1.70g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-30860MEDL

Quant Time: Aug 10 12:22:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	139192	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	248113	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	257593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	121884	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	108837	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
35) Dibromofluoromethane	7.55	113	77396	45.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.20%	
50) Toluene-d8	10.06	98	328175	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	12.38	95	137247	54.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.62%	
Target Compounds						
16) Acetone	3.78	43	20951	25.190	ug/l	95
39) Methylcyclohexane	9.05	83	5771	1.985	ug/l	91
52) Toluene	10.13	92	94594	21.252	ug/l	100
67) Ethyl Benzene	11.49	91	21727	2.255	ug/l #	89
68) m/p-Xylenes	11.59	106	13378	3.825	ug/l	87
69) o-Xylene	11.92	106	4595	1.402	ug/l	67
84) 1,2,4-Trimethylbenzene	13.02	105	16144	2.119	ug/l	88
89) n-Butylbenzene	13.59	91	18459	2.779	ug/l #	82
95) Naphthalene	15.10	128	229685	41.661	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081020\
Data File : VN062801.D
Acq On : 10 Aug 2020 11:47
Operator : JC/MD
Sample : L3563-01MEDL 10X
Misc : 1.70µ/5mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-30860MEDL

Quant Time: Aug 10 12:22:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

