

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081020\
 Data File : VN062813.D
 Acq On : 10 Aug 2020 16:59
 Operator : JC/MD
 Sample : L3618-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB179-GW-558-560

Quant Time: Aug 11 07:39:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	137330	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	238770	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	241146	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88977	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	112770	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
35) Dibromofluoromethane	7.55	113	81298	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
50) Toluene-d8	10.06	98	304741	47.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.58%	
62) 4-Bromofluorobenzene	12.38	95	116870	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
Target Compounds						
16) Acetone	3.79	43	4857	5.919	ug/l	98
52) Toluene	10.13	92	1632	0.381	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081020\
Data File : VN062813.D
Acq On : 10 Aug 2020 16:59
Operator : JC/MD
Sample : L3618-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB179-GW-558-560

Quant Time: Aug 11 07:39:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

