

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081020\
 Data File : VN062819.D
 Acq On : 10 Aug 2020 19:50
 Operator : JC/MD
 Sample : L3506-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-080720

Quant Time: Aug 11 07:56:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	142113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	252147	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	261062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	115409	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
35) Dibromofluoromethane	7.55	113	84114	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
50) Toluene-d8	10.06	98	332370	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
62) 4-Bromofluorobenzene	12.38	95	122433	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
Target Compounds						
16) Acetone	3.78	43	13036	15.351	ug/l	99
20) Methylene Chloride	4.50	84	13485	8.039	ug/l	89
43) Isopropyl Acetate	8.13	43	27995	6.544	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081020\
Data File : VN062819.D
Acq On : 10 Aug 2020 19:50
Operator : JC/MD
Sample : L3506-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-080720

Quant Time: Aug 11 07:56:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

