

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081021\
 Data File : VN068053.D
 Acq On : 10 Aug 2021 11:23
 Operator : JC/MD
 Sample : VN0810WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0810WBL01

Quant Time: Aug 11 06:30:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	321099	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	615563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	594969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	203978	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	246769	54.286	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.580%	
35) Dibromofluoromethane	8.024	113	191170	53.289	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.580%	
50) Toluene-d8	10.446	98	782328	47.465	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 94.940%	
62) 4-Bromofluorobenzene	12.734	95	251658	42.371	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 84.740%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081021\
Data File : VN068053.D
Acq On : 10 Aug 2021 11:23
Operator : JC/MD
Sample : VN0810WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0810WBL01

Quant Time: Aug 11 06:30:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

