

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081021\
 Data File : VN068056.D
 Acq On : 10 Aug 2021 13:03
 Operator : JC/MD
 Sample : M3318-05RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampled :
 SPLP-AMENDED-COM-9RE

Manual Integrations
APPROVED

MMDadoda
 8/11/2021 5:05:49 PM

Quant Time: Aug 11 06:31:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	336202	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	653038	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	619860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	210670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229891	48.301	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.600%
35) Dibromofluoromethane	8.024	113	158626	41.680	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	83.360%
50) Toluene-d8	10.446	98	731806	42.309	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	84.620%
62) 4-Bromofluorobenzene	12.736	95	237046	38.178	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	76.360%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	33499	19.290	ug/l	# 87
20) Methylene Chloride	5.098	84	12995m	2.973	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081021\
 Data File : VN068056.D
 Acq On : 10 Aug 2021 13:03
 Operator : JC/MD
 Sample : M3318-05RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampled :
 SPLP-AMENDED-COM-9RE

Manual Integrations
APPROVED
 MMDadoda
 8/11/2021 5:05:49 PM

Quant Time: Aug 11 06:31:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

