

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081021\
 Data File : VN068059.D
 Acq On : 10 Aug 2021 14:16
 Operator : JC/MD
 Sample : M3331-13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BSY1-WC-S-10-14-0-10

Manual Integrations
 APPROVED

MMDadoda
 8/11/2021 5:06:03 PM

Quant Time: Aug 11 06:31:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	312927	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	617244	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	647723	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	220700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	244488	55.189	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.380%
35) Dibromofluoromethane	8.024	113	187209	52.043	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.080%
50) Toluene-d8	10.443	98	793947	47.992	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.980%
62) 4-Bromofluorobenzene	12.733	95	270104	45.002	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.000%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	38532	23.838	ug/l	99
18) Methyl Acetate	4.867	43	4596m	1.082	ug/l	
20) Methylene Chloride	5.103	84	43416m	11.413	ug/l	
25) 2-Butanone	7.343	43	26903m	11.472	ug/l	
43) Isopropyl Acetate	8.568	43	35779	3.996	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081021\
Data File : VN068059.D
Acq On : 10 Aug 2021 14:16
Operator : JC/MD
Sample : M3331-13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BSY1-WC-S-10-14-0-10

Manual Integrations
APPROVED

MMDadoda
8/11/2021 5:06:03 PM

Quant Time: Aug 11 06:31:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

