

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081021\
 Data File : VN068060.D
 Acq On : 10 Aug 2021 15:01
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Aug 11 06:31:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	304242	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	582520	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	587637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	206379	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	233898	54.306	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.620%	
35) Dibromofluoromethane	8.024	113	180184	53.076	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.160%	
50) Toluene-d8	10.443	98	743846	47.672	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.340%	
62) 4-Bromofluorobenzene	12.736	95	250325	44.283	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.560%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081021\
Data File : VN068060.D
Acq On : 10 Aug 2021 15:01
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Aug 11 06:31:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

