

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
 Data File : VN073801.D
 Acq On : 10 Aug 2022 16:35
 Operator : JC\MD
 Sample : PB146860TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146860TB

Quant Time: Aug 10 22:55:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	395209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	635347	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	643313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	302113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	231881	48.145	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.300%
35) Dibromofluoromethane	7.951	113	217785	54.588	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.180%
50) Toluene-d8	10.380	98	746523	47.551	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.100%
62) 4-Bromofluorobenzene	12.674	95	281640	51.042	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.080%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	112830	43.617	ug/l	98
20) Methylene Chloride	5.010	84	33987	6.190	ug/l	98
25) 2-Butanone	7.275	43	22621	5.631	ug/l	92
37) Ethyl Acetate	7.340	43	64302	8.587	ug/l	99
43) Isopropyl Acetate	8.498	43	255427	25.254	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
Data File : VN073801.D
Acq On : 10 Aug 2022 16:35
Operator : JC\MD
Sample : PB146860TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146860TB

Quant Time: Aug 10 22:55:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

