

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
 Data File : VN073803.D
 Acq On : 10 Aug 2022 17:25
 Operator : JC\MD
 Sample : PB146860ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146860ZHE#02

Quant Time: Aug 10 22:55:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	383761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	634968	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	632098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	289286	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	229510	49.075	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.140%
35) Dibromofluoromethane	7.951	113	216659	54.339	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.680%
50) Toluene-d8	10.380	98	733076	46.722	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.440%
62) 4-Bromofluorobenzene	12.674	95	275004	49.932	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.860%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	83178	33.113	ug/l	99
18) Methyl Acetate	4.781	43	7781	1.245	ug/l	91
20) Methylene Chloride	5.016	84	14919	1.535	ug/l #	89
25) 2-Butanone	7.263	43	26936	6.905	ug/l	95
37) Ethyl Acetate	7.339	43	69331	9.264	ug/l	99
43) Isopropyl Acetate	8.498	43	272845	26.992	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
Data File : VN073803.D
Acq On : 10 Aug 2022 17:25
Operator : JC\MD
Sample : PB146860ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146860ZHE#02

Quant Time: Aug 10 22:55:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

