

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
 Data File : VN073805.D
 Acq On : 10 Aug 2022 18:14
 Operator : JC\MD
 Sample : PB146860ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146860ZHE#04

Quant Time: Aug 10 22:55:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	394174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	659592	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	649020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	303430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	230703	48.027	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.060%
35) Dibromofluoromethane	7.951	113	217611	52.540	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.080%
50) Toluene-d8	10.380	98	745913	45.766	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.540%
62) 4-Bromofluorobenzene	12.674	95	277233	48.539	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.080%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	80766	31.304	ug/l	99
18) Methyl Acetate	4.787	43	10079	1.570	ug/l	96
20) Methylene Chloride	5.016	84	15146	1.490	ug/l	97
25) 2-Butanone	7.269	43	28502	7.113	ug/l	90
37) Ethyl Acetate	7.339	43	69203	8.902	ug/l	99
43) Isopropyl Acetate	8.498	43	287281	27.359	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
Data File : VN073805.D
Acq On : 10 Aug 2022 18:14
Operator : JC\MD
Sample : PB146860ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146860ZHE#04

Quant Time: Aug 10 22:55:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

