

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
 Data File : VN073806.D
 Acq On : 10 Aug 2022 18:39
 Operator : JC\MD
 Sample : PB146860ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146860ZHE#05

Quant Time: Aug 10 22:55:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	373005	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	617678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	612177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	280111	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	223622	49.195	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.380%		
35) Dibromofluoromethane	7.951	113	211231	54.460	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	108.920%		
50) Toluene-d8	10.380	98	712248	46.665	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.340%		
62) 4-Bromofluorobenzene	12.674	95	262874	49.114	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.220%		
Target Compounds					Qvalue	
16) Acetone	4.228	43	83210	34.081	ug/l	98
18) Methyl Acetate	4.793	43	10858	1.788	ug/l #	74
20) Methylene Chloride	5.022	84	15873	1.898	ug/l #	85
25) 2-Butanone	7.269	43	25809	6.807	ug/l	95
37) Ethyl Acetate	7.345	43	70092	9.628	ug/l	99
43) Isopropyl Acetate	8.498	43	324277	32.978	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
Data File : VN073806.D
Acq On : 10 Aug 2022 18:39
Operator : JC\MD
Sample : PB146860ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146860ZHE#05

Quant Time: Aug 10 22:55:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

