

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
 Data File : VN073807.D
 Acq On : 10 Aug 2022 19:04
 Operator : JC\MD
 Sample : N4119-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP1-2-4FT

Quant Time: Aug 10 22:56:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	378209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	625295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	612885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	278830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	227303	49.316	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.640%
35) Dibromofluoromethane	7.951	113	215959	55.001	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.000%
50) Toluene-d8	10.380	98	726740	47.035	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.060%
62) 4-Bromofluorobenzene	12.674	95	263629	48.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.360%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	81577	32.953	ug/l	97
18) Methyl Acetate	4.781	43	9808	1.592	ug/l	90
20) Methylene Chloride	5.022	84	22741	3.635	ug/l	88
25) 2-Butanone	7.275	43	26222	6.821	ug/l	92
37) Ethyl Acetate	7.345	43	63847	8.663	ug/l	99
43) Isopropyl Acetate	8.498	43	254170	25.533	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
Data File : VN073807.D
Acq On : 10 Aug 2022 19:04
Operator : JC\MD
Sample : N4119-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP1-2-4FT

Quant Time: Aug 10 22:56:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

