

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
 Data File : VN073808.D
 Acq On : 10 Aug 2022 19:29
 Operator : JC\MD
 Sample : N4119-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP2-2FT

Quant Time: Aug 10 22:56:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	362330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	604112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	585621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	252431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	214966	48.684	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.360%
35) Dibromofluoromethane	7.951	113	205613	54.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.400%
50) Toluene-d8	10.380	98	687498	46.055	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.120%
62) 4-Bromofluorobenzene	12.674	95	247511	47.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.760%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	84648	35.692	ug/l	98
18) Methyl Acetate	4.781	43	8670	1.469	ug/l	94
20) Methylene Chloride	5.016	84	22459	3.818	ug/l	98
25) 2-Butanone	7.269	43	24773	6.726	ug/l	96
37) Ethyl Acetate	7.339	43	65732	9.232	ug/l	97
43) Isopropyl Acetate	8.498	43	244442	25.417	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
Data File : VN073808.D
Acq On : 10 Aug 2022 19:29
Operator : JC\MD
Sample : N4119-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP2-2FT

Quant Time: Aug 10 22:56:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

