

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
 Data File : VN073809.D
 Acq On : 10 Aug 2022 19:54
 Operator : JC\MD
 Sample : N4102-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FILTER-DRUM-0808

Quant Time: Aug 10 22:56:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	359659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	597884	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	586791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	283333	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	216140	49.313	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.620%
35) Dibromofluoromethane	7.951	113	204505	54.472	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.940%
50) Toluene-d8	10.380	98	685889	46.426	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.860%
62) 4-Bromofluorobenzene	12.674	95	258228	49.802	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.600%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	95958	40.761	ug/l	97
18) Methyl Acetate	4.787	43	9246	1.579	ug/l	95
20) Methylene Chloride	5.022	84	33332	6.850	ug/l #	89
25) 2-Butanone	7.269	43	31135	8.516	ug/l	100
37) Ethyl Acetate	7.339	43	67451	9.572	ug/l	98
43) Isopropyl Acetate	8.498	43	254281	26.715	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081022\
Data File : VN073809.D
Acq On : 10 Aug 2022 19:54
Operator : JC\MD
Sample : N4102-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FILTER-DRUM-0808

Quant Time: Aug 10 22:56:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

