

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081023\
 Data File : VN078830.D
 Acq On : 10 Aug 2023 12:19
 Operator : JC\MD
 Sample : PB154726ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154726ZHE#01

Quant Time: Aug 10 23:49:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080723W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 08 05:03:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	170140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	330352	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	304716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	79292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	135108	52.970	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.940%		
35) Dibromofluoromethane	8.177	113	112197	50.292	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.580%		
50) Toluene-d8	10.570	98	424694	51.731	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.460%		
62) 4-Bromofluorobenzene	12.853	95	111553	45.466	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 90.940%		
Target Compounds						
						Qvalue
16) Acetone	4.441	43	28589	29.131	ug/l	97
20) Methylene Chloride	5.288	84	12986	5.343	ug/l #	70
25) 2-Butanone	7.506	43	13594	8.180	ug/l #	64
37) Ethyl Acetate	7.577	43	5177	1.407	ug/l #	72
43) Isopropyl Acetate	8.700	43	119063	19.191	ug/l #	82
51) 4-Methyl-2-Pentanone	10.453	43	51128	13.454	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081023\
Data File : VN078830.D
Acq On : 10 Aug 2023 12:19
Operator : JC\MD
Sample : PB154726ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154726ZHE#01

Quant Time: Aug 10 23:49:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080723W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 08 05:03:52 2023
Response via : Initial Calibration

