

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081023\
 Data File : VN078839.D
 Acq On : 10 Aug 2023 15:56
 Operator : JC\MD
 Sample : PB154726ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154726ZHE#10

Quant Time: Aug 10 23:53:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080723W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 08 05:03:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	163678	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	331741	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	301719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	86947	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	134008	54.613	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.220%			
35) Dibromofluoromethane	8.177	113	109176	48.733	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.460%			
50) Toluene-d8	10.576	98	415461	50.394	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.780%			
62) 4-Bromofluorobenzene	12.859	95	113756	46.170	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.340%			
Target Compounds					Qvalue	
16) Acetone	4.447	43	61944	65.611	ug/l #	80
20) Methylene Chloride	5.289	84	15539	6.646	ug/l	89
25) 2-Butanone	7.506	43	30132	18.847	ug/l #	76
37) Ethyl Acetate	7.577	43	31247	8.456	ug/l #	92
43) Isopropyl Acetate	8.700	43	212898	34.171	ug/l #	71
51) 4-Methyl-2-Pentanone	10.453	43	115345	30.225	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081023\
Data File : VN078839.D
Acq On : 10 Aug 2023 15:56
Operator : JC\MD
Sample : PB154726ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154726ZHE#10

Quant Time: Aug 10 23:53:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080723W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 08 05:03:52 2023
Response via : Initial Calibration

