

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083202.D
 Acq On : 10 Aug 2024 12:56
 Operator : JC\MD
 Sample : PB162353TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162353TB

Quant Time: Aug 12 01:49:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	126834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	235765	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	239235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	104120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	104137	57.683	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.360%
35) Dibromofluoromethane	8.165	113	82452	56.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.060%
50) Toluene-d8	10.565	98	305911	55.728	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.460%
62) 4-Bromofluorobenzene	12.847	95	126007	58.879	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	117.760%
Target Compounds						
						Qvalue
20) Methylene Chloride	5.283	84	6257	3.849	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083202.D
Acq On : 10 Aug 2024 12:56
Operator : JC\MD
Sample : PB162353TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162353TB

Quant Time: Aug 12 01:49:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

