

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083205.D
 Acq On : 10 Aug 2024 14:08
 Operator : JC\MD
 Sample : P3476-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MUL04

Quant Time: Aug 12 01:51:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	126856	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	243733	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	247914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	109420	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	108271	59.962	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.920%
35) Dibromofluoromethane	8.171	113	83365	54.797	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.600%
50) Toluene-d8	10.565	98	317362	55.924	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.840%
62) 4-Bromofluorobenzene	12.847	95	130780	59.112	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	118.220%
Target Compounds						
						Qvalue
20) Methylene Chloride	5.271	84	7323	4.503	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083205.D
Acq On : 10 Aug 2024 14:08
Operator : JC\MD
Sample : P3476-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MUL04

Quant Time: Aug 12 01:51:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

