

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083207.D
 Acq On : 10 Aug 2024 14:57
 Operator : JC\MD
 Sample : P3476-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MUL06

Quant Time: Aug 12 02:44:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	142771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	277204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125979	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	118331	58.228	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 116.460%		
35) Dibromofluoromethane	8.171	113	94389	56.156	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 112.320%		
50) Toluene-d8	10.565	98	352974	56.297	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 112.600%		
62) 4-Bromofluorobenzene	12.847	95	148588	60.788	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 121.580%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	122304	153.805	ug/l	99
20) Methylene Chloride	5.283	84	15203	8.307	ug/l #	78
25) 2-Butanone	7.489	43	12725	10.422	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083207.D
Acq On : 10 Aug 2024 14:57
Operator : JC\MD
Sample : P3476-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MUL06

Quant Time: Aug 12 02:44:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

