

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083209.D
 Acq On : 10 Aug 2024 15:45
 Operator : JC\MD
 Sample : P3478-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TS04

Quant Time: Aug 12 01:52:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	143134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	274986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	287714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134488	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	122121	59.941	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 119.880%		
35) Dibromofluoromethane	8.165	113	95034	55.368	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 110.740%		
50) Toluene-d8	10.565	98	359932	56.217	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 112.440%		
62) 4-Bromofluorobenzene	12.847	95	156131	62.550	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 125.100%		
Target Compounds						
						Qvalue
16) Acetone	4.447	43	5232	6.563	ug/l #	82
20) Methylene Chloride	5.277	84	18593	10.134	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083209.D
Acq On : 10 Aug 2024 15:45
Operator : JC\MD
Sample : P3478-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TS04

Quant Time: Aug 12 01:52:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

