

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083211.D
 Acq On : 10 Aug 2024 16:34
 Operator : JC\MD
 Sample : P3478-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TS06

Quant Time: Aug 12 01:53:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	138898	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	263193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	267432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	116671	59.012	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.020%			
35) Dibromofluoromethane	8.165	113	90020	54.797	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.600%			
50) Toluene-d8	10.565	98	342263	55.852	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.700%			
62) 4-Bromofluorobenzene	12.847	95	141583	59.263	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 118.520%			
Target Compounds					Qvalue	
16) Acetone	4.436	43	3933	5.084	ug/l #	87
20) Methylene Chloride	5.277	84	16104	9.045	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083211.D
Acq On : 10 Aug 2024 16:34
Operator : JC\MD
Sample : P3478-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TS06

Quant Time: Aug 12 01:53:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

