

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083212.D
 Acq On : 10 Aug 2024 16:58
 Operator : JC\MD
 Sample : P3478-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TS07

Quant Time: Aug 12 01:54:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	137317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	270302	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118624	60.691	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.380%	
35) Dibromofluoromethane	8.171	113	92682	54.934	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.860%	
50) Toluene-d8	10.565	98	351135	55.793	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.580%	
62) 4-Bromofluorobenzene	12.847	95	149263	60.835	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 121.660%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.277	84	15580	8.851	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083212.D
Acq On : 10 Aug 2024 16:58
Operator : JC\MD
Sample : P3478-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TS07

Quant Time: Aug 12 01:54:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

