

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083213.D
 Acq On : 10 Aug 2024 17:22
 Operator : JC\MD
 Sample : P3479-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TS08

Quant Time: Aug 12 01:54:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	140723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	270671	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	129083	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	120249	60.033	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.060%	
35) Dibromofluoromethane	8.171	113	94137	55.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.440%	
50) Toluene-d8	10.565	98	352640	55.956	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.920%	
62) 4-Bromofluorobenzene	12.847	95	152295	61.986	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 123.980%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.277	84	13465	7.465	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083213.D
Acq On : 10 Aug 2024 17:22
Operator : JC\MD
Sample : P3479-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TS08

Quant Time: Aug 12 01:54:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

