

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083216.D
 Acq On : 10 Aug 2024 18:35
 Operator : JC\MD
 Sample : P3479-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TS11

Quant Time: Aug 12 01:56:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	134798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	261678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115905	60.408	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.820%
35) Dibromofluoromethane	8.171	113	91182	55.826	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.660%
50) Toluene-d8	10.571	98	340750	55.927	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.860%
62) 4-Bromofluorobenzene	12.847	95	142351	59.930	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	119.860%
Target Compounds						
						Qvalue
20) Methylene Chloride	5.277	84	16505	9.552	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083216.D
Acq On : 10 Aug 2024 18:35
Operator : JC\MD
Sample : P3479-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TS11

Quant Time: Aug 12 01:56:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

