

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083217.D
 Acq On : 10 Aug 2024 18:59
 Operator : JC\MD
 Sample : P3430-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 927-K1-WPO-0-0.25-073124

Quant Time: Aug 12 03:32:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	134468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257119	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	268284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	119104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115708	60.453	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.900%	
35) Dibromofluoromethane	8.171	113	88909	55.399	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.800%	
50) Toluene-d8	10.565	98	340326	56.848	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 113.700%#	
62) 4-Bromofluorobenzene	12.847	95	143039	61.287	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 122.580%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	2.512	62	6250	3.923	ug/l	98
16) Acetone	4.442	43	3688	4.924	ug/l	97
27) cis-1,2-Dichloroethene	7.489	96	2808	1.510	ug/l	38
29) Tetrahydrofuran	7.841	42	4401	5.920	ug/l #	90
44) Trichloroethene	9.359	130	501	0.291	ug/l	85
52) Toluene	10.629	92	2938	0.643	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083217.D
Acq On : 10 Aug 2024 18:59
Operator : JC\MD
Sample : P3430-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
927-K1-WPO-0-0.25-073124

Quant Time: Aug 12 03:32:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

