

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083219.D
 Acq On : 10 Aug 2024 19:47
 Operator : JC\MD
 Sample : P3506-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT4534

Quant Time: Aug 12 01:57:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	137825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262202	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125515	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116318	59.292	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 118.580%		
35) Dibromofluoromethane	8.171	113	91689	56.024	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 112.040%		
50) Toluene-d8	10.565	98	348513	57.087	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 114.180%#		
62) 4-Bromofluorobenzene	12.847	95	148821	62.528	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 125.060%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	25779	33.582	ug/l	90
20) Methylene Chloride	5.283	84	12255	6.937	ug/l #	84
25) 2-Butanone	7.482	43	17644	14.969	ug/l	92
43) Isopropyl Acetate	8.688	43	126578	25.090	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083219.D
Acq On : 10 Aug 2024 19:47
Operator : JC\MD
Sample : P3506-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT4534

Quant Time: Aug 12 01:57:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

