

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
 Data File : VN083220.D
 Acq On : 10 Aug 2024 20:11
 Operator : JC\MD
 Sample : P3530-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-S-04-202408

Quant Time: Aug 12 01:58:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	139535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	268829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	128211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	118866	59.848	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.700%
35) Dibromofluoromethane	8.165	113	92713	55.253	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.500%
50) Toluene-d8	10.565	98	356782	57.001	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	114.000%#
62) 4-Bromofluorobenzene	12.847	95	152090	62.327	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	124.660%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	24410	31.409	ug/l	97
20) Methylene Chloride	5.277	84	12181	6.810	ug/l	88
25) 2-Butanone	7.489	43	17125	14.351	ug/l	94
43) Isopropyl Acetate	8.688	43	137545	26.720	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081024\
Data File : VN083220.D
Acq On : 10 Aug 2024 20:11
Operator : JC\MD
Sample : P3530-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-S-04-202408

Quant Time: Aug 12 01:58:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

