

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081119\
 Data File : VN057186.D
 Acq On : 11 Aug 2019 14:39
 Operator : JC/SP
 Sample : K4227-27
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-K-(1)

Quant Time: Aug 12 05:55:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1373641	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1926894	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1586995	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	406659	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.44	65	629198	44.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.08%	
35) Dibromofluoromethane	6.99	113	574608	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
50) Toluene-d8	9.56	98	2241963	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
62) 4-Bromofluorobenzene	11.92	95	580661	38.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.92%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.15	74	2676	0.353	ug/l	97
11) Tert butyl alcohol	4.25	59	13001	6.467	ug/l #	73
16) Acetone	3.48	43	31270	7.329	ug/l	90
18) Methyl Acetate	3.90	43	29520	1.687	ug/l	94
20) Methylene Chloride	4.10	84	22276	1.582	ug/l	97
25) 2-Butanone	6.21	43	1569	0.241	ug/l #	47
29) Tetrahydrofuran	6.61	42	1091	0.248	ug/l #	70
31) Cyclohexane	7.05	56	20093	0.981	ug/l #	1
36) 1,1-Dichloropropene	7.05	75	81656	5.147	ug/l #	48
38) Carbon Tetrachloride	7.05	117	117799	7.125	ug/l #	17
43) Isopropyl Acetate	7.57	43	78305	3.509	ug/l #	75
45) 1,2-Dichloropropane	8.71	63	3221	0.248	ug/l #	43
48) Methyl methacrylate	8.62	41	18620	1.769	ug/l #	10
49) 1,4-Dioxane	8.70	88	379	1.768	ug/l #	1
51) 4-Methyl-2-Pentanone	9.49	43	18538	1.427	ug/l #	34
54) cis-1,3-Dichloropropene	9.36	75	130684	6.757	ug/l #	61
57) 1,3-Dichloropropane	10.25	76	23584	1.167	ug/l #	42
59) 2-Hexanone	10.25	43	325404	37.016	ug/l #	49
73) Isopropylbenzene	11.77	105	354	Below Cal	#	48
75) 1,1,2,2-Tetrachloroethane	11.92	83	336	Below Cal	#	1
76) 1,2,3-Trichloropropane	11.92	75	266083	32.485	ug/l #	100
79) 2-Chlorotoluene	12.20	91	472	Below Cal	#	63
81) trans-1,4-Dichloro-2-buten	11.92	75	266083	96.207	ug/l #	14
83) tert-Butylbenzene	12.52	119	158	Below Cal	#	66
93) 1,2,4-Trichlorobenzene	14.45	180	470	6.000	ug/l #	21
95) Naphthalene	14.67	128	1356	2.932	ug/l #	72
96) 1,2,3-Trichlorobenzene	14.83	180	221	5.753	ug/l #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN081119\
Data File : VN057186.D
Acq On : 11 Aug 2019 14:39
Operator : JC/SP
Sample : K4227-27
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-K-(1)

Quant Time: Aug 12 05:55:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.05 min Scan# 1642

Delta R.T. 0.00 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39

Instrument :

MSVOA_N

ClientSampleId :


1S-K-(1)

Tgt Ion:168 Resp: 1373641

Ion Ratio Lower Upper

168 100

99 48.5 38.6 57.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.05

600000

500000

400000



300000

200000

100000

0

Time--> 6.90 7.00 7.10 7.20



#8

Diethyl Ether

Concen: 0.353 ug/l

RT: 3.15 min Scan# 427

Delta R.T. 0.01 min

Lab File: VN057186.D


Acq: 11 Aug 2019 14:39

Tgt Ion: 74 Resp: 2676

Ion Ratio Lower Upper

74 100

45 95.0 46.3 138.8



Abundance Ion 74.00 (73.70 to 74.70): VN

Ion 45.00 (44.70 to 45.70): VN

3.15


1500


1000

500

0

Time--> 3.10 3.15 3.20





#11

Tert butyl alcohol

Concen: 6.467 ug/l

RT: 4.25 min Scan# 769

Delta R.T. 0.01 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39

Instrument :

MSVOA_N

ClientSampleId :





1S-K-(1)

Tgt Ion: 59 Resp: 13001

Ion Ratio Lower Upper

59 100

57 0.0 7.9 11.9#



#16

Acetone

Concen: 7.329 ug/l

RT: 3.48 min Scan# 531

Delta R.T. 0.01 min

Lab File: VN057186.D




Acq: 11 Aug 2019 14:39


Tgt Ion: 43 Resp: 31270

Ion Ratio Lower Upper

43 100

58 28.6 27.4 41.2









#18
Methyl Acetate
Concen: 1.687 ug/l
RT: 3.90 min Scan# 660
Delta R.T. 0.01 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39




Instrument :
MSVOA_N
ClientSampled :
1S-K-(1)


Tgt Ion: 43 Resp: 29520
Ion Ratio Lower Upper
43 100
74 22.3 20.2 30.4



#20
Methylene Chloride
Concen: 1.582 ug/l
RT: 4.10 min Scan# 724
Delta R.T. 0.00 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39

Tgt Ion: 84 Resp: 22276
Ion Ratio Lower Upper
84 100
49 109.7 89.5 134.3
51 37.3 27.2 40.8
86 64.8 50.6 75.8





#25

2-Butanone

Concen: 0.241 ug/l

RT: 6.21 min Scan# 1381

Delta R.T. 0.02 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39

Instrument :

MSVOA_N

ClientSampled :





1S-K-(1)

Tgt Ion: 43 Resp: 1569

Ion Ratio Lower Upper

43 100

72 0.0 22.2 33.4#



#29

Tetrahydrofuran

Concen: 0.248 ug/l

RT: 6.61 min Scan# 1503

Delta R.T. 0.01 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39




Tgt Ion: 42 Resp: 1091


Ion Ratio Lower Upper

42 100

72 18.7 37.1 55.7#

71 54.3 34.8 52.2#









#31
Cyclohexane
Concen: 0.981 ug/l
RT: 7.05 min Scan# 1641
Delta R.T. -0.02 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39

Instrument :
MSVOA_N
ClientSampleId :
1S-K-(1)

Tgt Ion: 56 Resp: 20093
Ion Ratio Lower Upper
56 100
69 189.3 26.9 40.3#
84 168.8 71.9 107.9#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 44.539 ug/l
RT: 7.44 min Scan# 1762
Delta R.T. -0.00 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39

Tgt Ion: 65 Resp: 629198
Ion Ratio Lower Upper
65 100
67 54.5 0.0 108.2



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.01 min Scan# 1939
 Delta R.T. -0.00 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-K-(1)

Tgt Ion:114 Resp: 1926894
 Ion Ratio Lower Upper
 114 100
 63 18.3 0.0 36.2
 88 15.0 0.0 30.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 47.522 ug/l
 RT: 6.99 min Scan# 1621
 Delta R.T. -0.00 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39

Tgt Ion:113 Resp: 574608
 Ion Ratio Lower Upper
 113 100
 111 103.2 81.3 121.9
 192 23.3 19.0 28.4



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 5.147 ug/l

RT: 7.05 min Scan# 1642

Delta R.T. -0.15 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39

Instrument :

MSVOA_N

ClientSampleId :

1S-K-(1)





Tgt Ion: 75 Resp: 81656

Ion Ratio Lower Upper

75 100

110 8.7 18.7 56.1#

77 0.0 25.0 37.4#



#38

Carbon Tetrachloride

Concen: 7.125 ug/l

RT: 7.05 min Scan# 1641

Delta R.T. -0.14 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39




Tgt Ion: 117 Resp: 117799


Ion Ratio Lower Upper

117 100

119 5.8 75.8 113.6#

121 0.0 24.4 36.6#





#43

Isopropyl Acetate

Concen: 3.509 ug/l

RT: 7.57 min Scan# 1803

Delta R.T. 0.01 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39

Instrument :

MSVOA_N

ClientSampleId :

1S-K-(1)


Tgt Ion: 43 Resp: 78305

Ion Ratio Lower Upper

43 100

61 14.0 24.2 36.2#




87 8.4 11.8 17.8#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#45

1,2-Dichloropropane

Concen: 0.248 ug/l

RT: 8.71 min Scan# 2156

Delta R.T. 0.14 min

Lab File: VN057186.D


Acq: 11 Aug 2019 14:39

Tgt Ion: 63 Resp: 3221

Ion Ratio Lower Upper



63 100


65 0.0 25.2 37.8#



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 64.90 (64.60 to 65.60): VN0









#48
Methyl methacrylate
Concen: 1.769 ug/l
RT: 8.62 min Scan# 2129
Delta R.T. -0.02 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39




Instrument :
MSVOA_N
ClientSampleId :
1S-K-(1)


Tgt Ion: 41 Resp: 18620
Ion Ratio Lower Upper
41 100
69 0.0 77.5 116.3#
39 0.0 39.8 59.6#



#49
1,4-Dioxane
Concen: 1.768 ug/l
RT: 8.70 min Scan# 2155
Delta R.T. 0.06 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39

Tgt Ion: 88 Resp: 379
Ion Ratio Lower Upper
88 100
43 743058.6 21.7 32.5#
58 3611.3 51.1 76.7#









#50
Toluene-d8
Concen: 48.464 ug/l
RT: 9.56 min Scan# 2421
Delta R.T. -0.00 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39




Instrument :
MSVOA_N
ClientSampled :
1S-K-(1)


Tgt Ion: 98 Resp: 2241963
Ion Ratio Lower Upper
98 100
100 64.2 51.2 76.8



#51
4-Methyl-2-Pentanone
Concen: 1.427 ug/l
RT: 9.49 min Scan# 2399
Delta R.T. 0.04 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39

Tgt Ion: 43 Resp: 18538
Ion Ratio Lower Upper
43 100
58 0.0 33.2 49.8#









#54

cis-1,3-Dichloropropene
 Concen: 6.757 ug/l
 RT: 9.36 min Scan# 2361
 Delta R.T. 0.07 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-K-(1)




Tgt Ion	Ratio	Lower	Upper
75	100		
77	0.7	25.1	37.7#
39	24.4	32.5	48.7#




#57

1,3-Dichloropropane
 Concen: 1.167 ug/l
 RT: 10.25 min Scan# 2636
 Delta R.T. 0.06 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39

Tgt Ion	Ratio	Lower	Upper
76	100		
78	0.0	25.9	38.9#









#59
2-Hexanone
Concen: 37.016 ug/l
RT: 10.25 min Scan# 2636
Delta R.T. 0.01 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39




Instrument :
MSVOA_N
ClientSampled :
1S-K-(1)


Tgt Ion: 43 Resp: 325404
Ion Ratio Lower Upper
43 100
58 20.4 29.2 87.6#



#62
4-Bromofluorobenzene
Concen: 38.956 ug/l
RT: 11.92 min Scan# 3156
Delta R.T. 0.00 min
Lab File: VN057186.D
Acq: 11 Aug 2019 14:39

Tgt Ion: 95 Resp: 580661
Ion Ratio Lower Upper
95 100
174 94.8 0.0 185.0
176 92.3 0.0 181.2









#63
 Chlorobenzene-d5
 Concen: 50.000 ug/l
 RT: 10.91 min Scan# 2841
 Delta R.T. 0.00 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39




Instrument :
 MSVOA_N
 ClientSampled :
 1S-K-(1)


Tgt Ion:117 Resp: 1586995
 Ion Ratio Lower Upper
 117 100
 82 50.9 40.8 61.2
 119 32.0 25.7 38.5



#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.88 min Scan# 3453
 Delta R.T. 0.00 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39

Tgt Ion:152 Resp: 406659
 Ion Ratio Lower Upper
 152 100
 115 53.8 27.5 82.5
 150 156.2 0.0 341.6





#73

Isopropylbenzene

Concen: Below Cal

RT: 11.77 min Scan# 3108

Delta R.T. 0.00 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39

Instrument :

MSVOA_N

ClientSampleId :


1S-K-(1)

Tgt Ion: 105 Resp: 354

Ion Ratio Lower Upper

105 100

120 0.0 13.7 41.0#



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

11.77

400



300

200

100

0

Time-->



#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 11.92 min Scan# 3157

Delta R.T. -0.10 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39


Tgt Ion: 83 Resp: 336

Ion Ratio Lower Upper

83 100

131 239.0 5.5 16.4#

85 0.0 32.4 97.0#



Abundance Ion 82.90 (82.60 to 83.60): V

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): V

600

500

400


300


200

100

0

Time-->





#76

1,2,3-Trichloropropane

Concen: 32.485 ug/l

RT: 11.92 min Scan# 3155

Delta R.T. -0.16 min

Lab File: VN057186.D

Acq: 11 Aug 2019 14:39

Instrument :

MSVOA_N

ClientSampleId :


1S-K-(1)

Tgt Ion: 75 Resp: 266083

Ion Ratio Lower Upper




75 100

77 0.5 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0



#79

2-Chlorotoluene

Concen: Below Cal

RT: 12.20 min Scan# 3244

Delta R.T. 0.01 min

Lab File: VN057186.D


Acq: 11 Aug 2019 14:39

Tgt Ion: 91 Resp: 472

Ion Ratio Lower Upper



91 100


126 14.6 18.1 54.3#



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 125.90 (125.60 to 126.60): V









#81
 trans-1,4-Dichloro-2-butene
 Concen: 96.207 ug/l
 RT: 11.92 min Scan# 3155
 Delta R.T. 0.10 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39




Instrument :
 MSVOA_N
 ClientSampleId :
 1S-K-(1)


Tgt Ion: 75 Resp: 266083
 Ion Ratio Lower Upper
 75 100
 53 0.0 70.6 106.0#
 89 0.0 36.4 54.6#



#83
 tert-Butylbenzene
 Concen: Below Cal
 RT: 12.52 min Scan# 3342
 Delta R.T. -0.00 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39

Tgt Ion: 119 Resp: 158
 Ion Ratio Lower Upper
 119 100
 91 42.4 31.4 94.2
 134 0.0 13.2 39.6#









#93
 1,2,4-Trichlorobenzene
 Concen: 6.000 ug/l
 RT: 14.45 min Scan# 3944
 Delta R.T. 0.01 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39




Instrument :
 MSVOA_N
 ClientSampled :
 1S-K-(1)


Tgt Ion:180 Resp: 470
 Ion Ratio Lower Upper
 180 100
 182 7.0 47.5 142.6#
 145 6.6 14.1 42.2#



#95
 Naphthalene
 Concen: 2.932 ug/l
 RT: 14.67 min Scan# 4010
 Delta R.T. 0.00 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39

Tgt Ion:128 Resp: 1356
 Ion Ratio Lower Upper
 128 100
 127 0.0 10.3 15.5#
 129 2.2 8.8 13.2#





#96
 1,2,3-Trichlorobenzene
 Concen: 5.753 ug/l
 RT: 14.83 min Scan# 4061
 Delta R.T. -0.00 min
 Lab File: VN057186.D
 Acq: 11 Aug 2019 14:39

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-K-(1)

Tgt Ion	Ratio	Lower	Upper
180	100		
182	227.1	48.3	144.8#
145	0.0	14.9	44.5#

