

Data File : VN057177.D
Acq On : 11 Aug 2019 11:01
Operator : JC/SP
Sample : VN0811WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0811WBL01

Quant Time: Aug 12 06:25:19 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1378372	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1920819	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.90	117	1589696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	404789	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	633914	44.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.44%	
35) Dibromofluoromethane	6.99	113	580143	48.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.26%	
50) Toluene-d8	9.56	98	2201457	47.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.48%	
62) 4-Bromofluorobenzene	11.92	95	583302	39.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	78.52%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057177.D
Acq On : 11 Aug 2019 11:01
Operator : JC/SP
Sample : VN0811WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0811WBL01

Quant Time: Aug 12 06:25:19 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

