

Data File : VN057182.D
Acq On : 11 Aug 2019 13:02
Operator : JC/SP
Sample : K4227-23
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-I(1)

Quant Time: Aug 12 06:42:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1396246	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1949249	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1596101	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	416099	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	641793	44.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.40%	
35) Dibromofluoromethane	6.99	113	586694	47.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.92%	
50) Toluene-d8	9.56	98	2267218	48.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.90%	
62) 4-Bromofluorobenzene	11.92	95	605637	40.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.34%	
Target Compounds						
16) Acetone	3.48	43	31782	7.329	ug/l	89
20) Methylene Chloride	4.10	84	19124	1.336	ug/l	92
43) Isopropyl Acetate	7.56	43	47754	2.115	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057182.D
Acq On : 11 Aug 2019 13:02
Operator : JC/SP
Sample : K4227-23
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-I(1)

Quant Time: Aug 12 06:42:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

