

Data File : VN057183.D
Acq On : 11 Aug 2019 13:26
Operator : JC/SP
Sample : K4227-24
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-J(0)

Quant Time: Aug 12 06:49:34 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1388542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1953769	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1642938	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	429834	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	640506	44.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.70%	
35) Dibromofluoromethane	6.99	113	588611	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
50) Toluene-d8	9.56	98	2268880	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
62) 4-Bromofluorobenzene	11.92	95	610145	40.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.74%	
Target Compounds						
16) Acetone	3.48	43	23723	5.501	ug/l	100
20) Methylene Chloride	4.10	84	15677	1.101	ug/l #	89
43) Isopropyl Acetate	7.57	43	62989	2.783	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057183.D
Acq On : 11 Aug 2019 13:26
Operator : JC/SP
Sample : K4227-24
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-J(0)

Quant Time: Aug 12 06:49:34 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

