

Data File : VN057186.D
Acq On : 11 Aug 2019 14:39
Operator : JC/SP
Sample : K4227-27
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-K-(1)

Quant Time: Aug 12 06:59:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1373641	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1926894	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1586995	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	406659	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	629198	44.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.08%	
35) Dibromofluoromethane	6.99	113	574608	47.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.04%	
50) Toluene-d8	9.56	98	2241963	48.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.92%	
62) 4-Bromofluorobenzene	11.92	95	580661	38.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.92%	
Target Compounds						
16) Acetone	3.48	43	31270	7.329	ug/l	90
18) Methyl Acetate	3.90	43	29520	1.687	ug/l	94
20) Methylene Chloride	4.10	84	22276	1.582	ug/l	97
43) Isopropyl Acetate	7.57	43	78305	3.509	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057186.D
Acq On : 11 Aug 2019 14:39
Operator : JC/SP
Sample : K4227-27
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-K-(1)

Quant Time: Aug 12 06:59:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

