

Data File : VN057187.D
Acq On : 11 Aug 2019 15:03
Operator : JC/SP
Sample : K4227-28
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-P-(0)

Quant Time: Aug 12 07:01:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1357567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1909899	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1590227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	416300	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	626924	44.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.80%	
35) Dibromofluoromethane	6.99	113	570867	47.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.26%	
50) Toluene-d8	9.56	98	2237452	48.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.92	95	586974	39.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.46%	
Target Compounds						
16) Acetone	3.48	43	33025	7.832	ug/l	100
18) Methyl Acetate	3.90	43	29095	1.681	ug/l	99
20) Methylene Chloride	4.11	84	23192	1.667	ug/l #	91
43) Isopropyl Acetate	7.56	43	86715	3.920	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057187.D
Acq On : 11 Aug 2019 15:03
Operator : JC/SP
Sample : K4227-28
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-P-(0)

Quant Time: Aug 12 07:01:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

