

Data File : VN057189.D
Acq On : 11 Aug 2019 15:52
Operator : JC/SP
Sample : K4227-30
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-Q(0)

Manual Integrations
APPROVED

MMDadoda
8/12/2019 3:41:41 PM

Quant Time: Aug 12 07:05:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1403309	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1964272	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1620951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	420873	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	640197	44.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.72%	
35) Dibromofluoromethane	6.99	113	589668	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
50) Toluene-d8	9.56	98	2287235	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
62) 4-Bromofluorobenzene	11.92	95	597932	39.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.70%	
Target Compounds						
16) Acetone	3.48	43	25668	5.889	ug/l	93
18) Methyl Acetate	3.90	43	41139	2.544	ug/l	99
20) Methylene Chloride	4.10	84	18730	1.302	ug/l	97
43) Isopropyl Acetate	7.57	43	145040m	6.375	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057189.D
Acq On : 11 Aug 2019 15:52
Operator : JC/SP
Sample : K4227-30
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-Q-(0)

Manual Integrations
APPROVED

MMDadoda
8/12/2019 3:41:41 PM

Quant Time: Aug 12 07:05:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

