

Data File : VN057190.D
Acq On : 11 Aug 2019 16:16
Operator : JC/SP
Sample : K4227-31
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-Q(1)

Quant Time: Aug 12 07:07:17 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1359128	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1931245	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1629020	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	438028	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	638411	45.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.34%	
35) Dibromofluoromethane	6.99	113	574996	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
50) Toluene-d8	9.56	98	2258983	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
62) 4-Bromofluorobenzene	11.92	95	601855	40.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.58%	
Target Compounds						
16) Acetone	3.48	43	26774	6.343	ug/l	88
18) Methyl Acetate	3.90	43	39972	2.555	ug/l	98
20) Methylene Chloride	4.10	84	16333	1.172	ug/l	97
43) Isopropyl Acetate	7.57	43	191515	8.562	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057190.D
Acq On : 11 Aug 2019 16:16
Operator : JC/SP
Sample : K4227-31
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-Q(1)

Quant Time: Aug 12 07:07:17 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

