

Data File : VN057191.D
Acq On : 11 Aug 2019 16:41
Operator : JC/SP
Sample : K4227-32
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-O-(0)

Quant Time: Aug 12 07:09:02 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1385932	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1948625	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1628918	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	408295	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	648512	45.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.00%	
35) Dibromofluoromethane	6.98	113	591714	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
50) Toluene-d8	9.56	98	2278302	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
62) 4-Bromofluorobenzene	11.92	95	587860	39.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.00%	
Target Compounds						
16) Acetone	3.48	43	24660	5.729	ug/l	100
18) Methyl Acetate	3.90	43	52890	3.513	ug/l	100
20) Methylene Chloride	4.10	84	18706	1.317	ug/l #	84
43) Isopropyl Acetate	7.57	43	172926	7.662	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057191.D
Acq On : 11 Aug 2019 16:41
Operator : JC/SP
Sample : K4227-32
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-O-(0)

Quant Time: Aug 12 07:09:02 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

