

Data File : VN057196.D
Acq On : 11 Aug 2019 18:42
Operator : JC/SP
Sample : K4245-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-27

Quant Time: Aug 12 07:22:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1310085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1854699	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1548124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	409253	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	617336	45.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.64%	
35) Dibromofluoromethane	6.99	113	557173	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
50) Toluene-d8	9.56	98	2164190	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.92	95	576268	40.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.34%	
Target Compounds						
16) Acetone	3.48	43	39329	9.666	ug/l	99
18) Methyl Acetate	3.90	43	44514	3.055	ug/l	99
20) Methylene Chloride	4.10	84	103944	7.740	ug/l	97
43) Isopropyl Acetate	7.57	43	130065	6.055	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057196.D
Acq On : 11 Aug 2019 18:42
Operator : JC/SP
Sample : K4245-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-27

Quant Time: Aug 12 07:22:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

