

Data File : VN057198.D
Acq On : 11 Aug 2019 19:31
Operator : JC/SP
Sample : K4244-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BERGEN-SP-2

Manual Integrations
APPROVED

MMDadoda
8/12/2019 3:41:51 PM

Quant Time: Aug 12 07:27:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1311232	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1858372	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1576820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	425591	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	617838	45.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.64%	
35) Dibromofluoromethane	6.99	113	560693	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
50) Toluene-d8	9.56	98	2177338	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.92	95	589463	41.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.02%	
Target Compounds						
16) Acetone	3.48	43	28318	6.953	ug/l	99
18) Methyl Acetate	3.90	43	47513	3.302	ug/l	96
20) Methylene Chloride	4.10	84	16933	1.260	ug/l	95
43) Isopropyl Acetate	7.57	43	134931m	6.269	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN057198.D
Acq On : 11 Aug 2019 19:31
Operator : JC/SP
Sample : K4244-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BERGEN-SP-2

Manual Integrations
APPROVED

MMDadoda
8/12/2019 3:41:51 PM

Quant Time: Aug 12 07:27:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

