

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081120\
 Data File : VN062828.D
 Acq On : 11 Aug 2020 14:17
 Operator : JC/MD
 Sample : L3624-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS-TP-2

Quant Time: Aug 11 15:09:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	125045	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	220506	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	216839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	81429	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105386	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
35) Dibromofluoromethane	7.55	113	74451	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	10.06	98	282793	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
62) 4-Bromofluorobenzene	12.38	95	103779	46.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.26%	
Target Compounds						
16) Acetone	3.79	43	6528	8.737	ug/l	98
20) Methylene Chloride	4.50	84	7707	4.951	ug/l #	83
30) Chloroform	7.33	83	5929	2.043	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081120\
Data File : VN062828.D
Acq On : 11 Aug 2020 14:17
Operator : JC/MD
Sample : L3624-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS-TP-2

Quant Time: Aug 11 15:09:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

