

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081120\
 Data File : VN062832.D
 Acq On : 11 Aug 2020 15:54
 Operator : JC/MD
 Sample : PB130875ZHE#17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130875ZHE#17

Quant Time: Aug 11 21:32:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	123789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	216965	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	208676	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	82402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	107736	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
35) Dibromofluoromethane	7.55	113	76263	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
50) Toluene-d8	10.06	98	266578	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
62) 4-Bromofluorobenzene	12.38	95	104073	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
Target Compounds						
20) Methylene Chloride	4.49	84	5500	3.354	ug/l #	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081120\
Data File : VN062832.D
Acq On : 11 Aug 2020 15:54
Operator : JC/MD
Sample : PB130875ZHE#17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130875ZHE#17

Quant Time: Aug 11 21:32:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

