

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081120\
 Data File : VN062835.D
 Acq On : 11 Aug 2020 17:31
 Operator : JC/MD
 Sample : L3624-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS-TP-1

Quant Time: Aug 11 21:36:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	119448	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	207407	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	222112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	79514	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	101627	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
35) Dibromofluoromethane	7.55	113	71781	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
50) Toluene-d8	10.06	98	280216	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
62) 4-Bromofluorobenzene	12.38	95	102517	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
Target Compounds						
16) Acetone	3.78	43	17038	23.871	ug/l #	88
20) Methylene Chloride	4.50	84	7851	5.331	ug/l	90
43) Isopropyl Acetate	8.13	43	37565	10.676	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081120\
Data File : VN062835.D
Acq On : 11 Aug 2020 17:31
Operator : JC/MD
Sample : L3624-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS-TP-1

Quant Time: Aug 11 21:36:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

