

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081120\
 Data File : VN062836.D
 Acq On : 11 Aug 2020 17:56
 Operator : JC/MD
 Sample : L3624-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS-TP-2

Quant Time: Aug 11 21:38:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	122929	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	214013	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	227866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	83954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	104026	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
35) Dibromofluoromethane	7.55	113	74551	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
50) Toluene-d8	10.06	98	287599	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	12.38	95	106062	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
Target Compounds						
16) Acetone	3.78	43	19881	27.066	ug/l	94
20) Methylene Chloride	4.49	84	8223m	5.440	ug/l	
43) Isopropyl Acetate	8.13	43	37649	10.369	ug/l	98
51) 4-Methyl-2-Pentanone	9.96	43	4727	2.012	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081120\
Data File : VN062836.D
Acq On : 11 Aug 2020 17:56
Operator : JC/MD
Sample : L3624-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS-TP-2

Quant Time: Aug 11 21:38:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

