

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081120\
 Data File : VN062837.D
 Acq On : 11 Aug 2020 18:20
 Operator : JC/MD
 Sample : L3624-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS-TP-3

Quant Time: Aug 11 21:40:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	123564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	214272	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	227520	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	81331	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105866	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
35) Dibromofluoromethane	7.55	113	75073	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
50) Toluene-d8	10.06	98	287102	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
62) 4-Bromofluorobenzene	12.38	95	106417	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
Target Compounds						
16) Acetone	3.79	43	16585	22.462	ug/l	93
20) Methylene Chloride	4.51	84	8236	5.417	ug/l	94
43) Isopropyl Acetate	8.13	43	36215	9.962	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081120\
Data File : VN062837.D
Acq On : 11 Aug 2020 18:20
Operator : JC/MD
Sample : L3624-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS-TP-3

Quant Time: Aug 11 21:40:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

