

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081120\
 Data File : VN062839.D
 Acq On : 11 Aug 2020 19:09
 Operator : JC/MD
 Sample : PB130874ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130874ZHE#01

Quant Time: Aug 11 21:43:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	125588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	220822	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	234937	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	84546	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	107947	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
35) Dibromofluoromethane	7.55	113	75480	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
50) Toluene-d8	10.06	98	294849	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	12.38	95	107181	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
Target Compounds						
16) Acetone	3.78	43	13431	17.898	ug/l	91
20) Methylene Chloride	4.51	84	7897	5.067	ug/l	92
43) Isopropyl Acetate	8.13	43	39562	10.560	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081120\
Data File : VN062839.D
Acq On : 11 Aug 2020 19:09
Operator : JC/MD
Sample : PB130874ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130874ZHE#01

Quant Time: Aug 11 21:43:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

