

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081120\
 Data File : VN062845.D
 Acq On : 11 Aug 2020 21:34
 Operator : JC/MD
 Sample : L3607-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ES-TOTE-COMP

Quant Time: Aug 11 22:36:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	118731	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	214283	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	212987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80217	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	102757	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
35) Dibromofluoromethane	7.55	113	74258	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
50) Toluene-d8	10.06	98	271995	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
62) 4-Bromofluorobenzene	12.38	95	99414	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
Target Compounds						
16) Acetone	3.78	43	10957	15.444	ug/l	98
20) Methylene Chloride	4.50	84	2334m	1.053	ug/l	
52) Toluene	10.13	92	221821	57.703	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081120\
Data File : VN062845.D
Acq On : 11 Aug 2020 21:34
Operator : JC/MD
Sample : L3607-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ES-TOTE-COMP

Quant Time: Aug 11 22:36:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

