

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
 Data File : VN068065.D
 Acq On : 11 Aug 2021 11:47
 Operator : JC/MD
 Sample : VN0811MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0811MBL01

Quant Time: Aug 12 07:10:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	297543	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	581337	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	559331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	181364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	245306	58.237	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 116.480%	
35) Dibromofluoromethane	8.021	113	190063	56.100	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 112.200%	
50) Toluene-d8	10.443	98	778583	49.811	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.620%	
62) 4-Bromofluorobenzene	12.736	95	242729	43.167	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 86.340%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
Data File : VN068065.D
Acq On : 11 Aug 2021 11:47
Operator : JC/MD
Sample : VN0811MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0811MBL01

Quant Time: Aug 12 07:10:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

