

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
 Data File : VN068071.D
 Acq On : 11 Aug 2021 14:30
 Operator : JC/MD
 Sample : PB138317ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138317ZHE#01

Quant Time: Aug 12 07:12:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	296515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	575229	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	633534	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	213379	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	232649	55.423	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.840%
35) Dibromofluoromethane	8.024	113	180878	53.956	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.920%
50) Toluene-d8	10.443	98	769788	49.775	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.540%
62) 4-Bromofluorobenzene	12.734	95	261442	46.549	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.100%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	32998	21.544	ug/l	98
20) Methylene Chloride	5.093	84	16890	4.517	ug/l	89
43) Isopropyl Acetate	8.566	43	38332	4.594	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
Data File : VN068071.D
Acq On : 11 Aug 2021 14:30
Operator : JC/MD
Sample : PB138317ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138317ZHE#01

Quant Time: Aug 12 07:12:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

