

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
 Data File : VN068072.D
 Acq On : 11 Aug 2021 14:55
 Operator : JC/MD
 Sample : PB138317ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138317ZHE#02

Manual Integrations
 APPROVED

MMDadoda
 8/12/2021 4:51:14 PM

Quant Time: Aug 12 07:12:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.085	168	295471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	580133	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	605531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	191484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	232045	55.475	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 110.940%			
35) Dibromofluoromethane	8.024	113	175691	51.965	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 103.940%			
50) Toluene-d8	10.443	98	755973	48.570	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.140%			
62) 4-Bromofluorobenzene	12.733	95	241978	43.128	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 86.260%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	34067	22.321	ug/l	99
18) Methyl Acetate	4.859	43	5885m	1.504	ug/l	
20) Methylene Chloride	5.092	84	17629	4.745	ug/l	86
43) Isopropyl Acetate	8.566	43	42253	5.021	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081121\
Data File : VN068072.D
Acq On : 11 Aug 2021 14:55
Operator : JC/MD
Sample : PB138317ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138317ZHE#02

Manual Integrations
APPROVED

MMDadoda
8/12/2021 4:51:14 PM

Quant Time: Aug 12 07:12:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

